

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL111220\
 Data File : PL062991.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Nov 2020 15:50
 Operator : DD\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 13 00:31:13 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102120.M
 Quant Title : GC Extractables
 QLast Update : Thu Oct 22 04:25:31 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.411	3.951	85753102	156.8E6	49.854	53.373
28)	SA Decachlor...	8.126	8.991	76593439	99532827	47.743	47.556
Target Compounds							
2)	A alpha-BHC	3.843	4.337	125.9E6	235.6E6	49.361	53.029
3)	MA gamma-BHC...	4.122	4.620	113.2E6	218.3E6	48.884	52.280
4)	MA Heptachlor	4.419	5.126	103.0E6	192.0E6	47.978	48.157
5)	MB Aldrin	4.664	5.429	94378586	181.3E6	49.445	50.083
6)	B beta-BHC	4.372	4.792	47236044	94101692	49.452	49.974
7)	B delta-BHC	4.572	5.007	107.7E6	199.8E6	49.536	57.080
8)	B Heptachlo...	5.102	5.814	90749062	164.4E6	49.737	50.754
9)	A Endosulfan I	5.437	6.170	84219571	148.3E6	50.382	47.023
10)	B gamma-Chl...	5.328	6.048	89803119	161.2E6	49.847	47.874
11)	B alpha-Chl...	5.385	6.121	89888837	158.9E6	49.724	47.181
12)	B 4,4'-DDE	5.550	6.279	81831558	152.0E6	49.462	47.179
13)	MA Dieldrin	5.678	6.426	86786948	149.3E6	49.116	46.647
14)	MA Endrin	5.933	6.643	72527832	117.2E6	43.592	42.001
15)	B Endosulfa...	6.206	6.853	79045125	124.3E6	47.946	43.476
16)	A 4,4'-DDD	6.063	6.769	73935913	123.0E6	52.270	48.062
17)	MA 4,4'-DDT	6.301	7.067	57624030	94622655	43.178	42.143
18)	B Endrin al...	6.373	6.977	68110517	107.2E6	46.905	46.991
19)	B Endosulfa...	6.585	7.200	78103189	113.9E6	47.328	47.485
20)	A Methoxychlor	6.846	7.527	35945340	50455000	41.734	40.838
21)	B Endrin ke...	7.071	7.669	91971964	121.2E6	51.010	51.156
22)	Mirex	7.258	8.122	66921167	79393686	47.204	46.590

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL111220\
Data File : PL062991.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2020 15:50
Operator : DD\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 13 00:31:13 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102120.M
Quant Title : GC Extractables
QLast Update : Thu Oct 22 04:25:31 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

