

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL111324\
 Data File : PL093046.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Nov 2024 18:23
 Operator : AR\AJ
 Sample : P4799-07
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 WC-TA2-03-C

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 14 00:38:18 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102824.M
 Quant Title : GC Extractables
 QLast Update : Mon Oct 28 18:58:23 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.538	2.776	51670629	57957718	21.088	21.299
28)	SA Decachlor...	9.051	7.913	35955204	54137932	18.682	19.836
Target Compounds							
2)	A alpha-BHC	0.000	3.266	0	2389013	N.D.	0.599 #
4)	MA Heptachlor	0.000	3.942	0	4422835	N.D.	1.172 #
5)	MB Aldrin	0.000	4.216	0	1134893	N.D.	0.310 #
7)	B delta-BHC	4.766	4.121f	4995261	992482	1.596	0.258 #
9)	A Endosulfan I	0.000	5.129f	0	311790	N.D.	0.102 #
10)	B gamma-Chl...	0.000	4.977	0	2796866	N.D.	0.832 #
11)	B alpha-Chl...	0.000	5.021f	0	4067813	N.D.	1.223 #
13)	MA Dieldrin	0.000	5.352	0	1679191	N.D.	0.503 #
14)	MA Endrin	0.000	5.655	0	1457763	N.D.	0.504 #
16)	A 4,4'-DDD	0.000	5.771f	0	548435	N.D.	0.220 #
17)	MA 4,4'-DDT	0.000	6.030	0	660611	N.D.	0.246 #
20)	A Methoxychlor	0.000	6.596f	0	449491	N.D.	0.315 #
25)	Chlordane-3	0.000	4.977	0	2796866	N.D.	7.712 #
26)	Chlordane-4	0.000	5.021f	0	4067813	N.D.	11.629 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL111324\
Data File : PL093046.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2024 18:23
Operator : AR\AJ
Sample : P4799-07
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
WC-TA2-03-C

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 00:38:18 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102824.M
Quant Title : GC Extractables
QLast Update : Mon Oct 28 18:58:23 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.538 min
Delta R.T.: -0.002 min
Response: 51670629
Conc: 21.09 ng/ml

Instrument :
ECD_L
ClientSampleId :
WC-TA2-03-C




#1 Tetrachloro-m-xylene

R.T.: 2.776 min
Delta R.T.: -0.002 min
Response: 57957718
Conc: 21.30 ng/ml




#2 alpha-BHC

R.T.: 0.000 min
Exp R.T. : 3.996 min
Response: 0
Conc: N.D.



#2 alpha-BHC


R.T.: 3.266 min
Delta R.T.: -0.015 min
Response: 2389013
Conc: 0.60 ng/ml



#4 Heptachlor


R.T.: 0.000 min
Exp R.T.: 4.917 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
WC-TA2-03-C




#4 Heptachlor

R.T.: 3.942 min
Delta R.T.: -0.008 min
Response: 4422835
Conc: 1.17 ng/ml




#5 Aldrin

R.T.: 0.000 min
Exp R.T.: 5.258 min
Response: 0
Conc: N.D.



#5 Aldrin


R.T.: 4.216 min
Delta R.T.: -0.014 min
Response: 1134893
Conc: 0.31 ng/ml



#7 delta-BHC


R.T.: 4.766 min
Delta R.T.: -0.006 min
Response: 4995261
Conc: 1.60 ng/ml

Instrument :
ECD_L
ClientSampleId :
WC-TA2-03-C




#7 delta-BHC

R.T.: 4.121 min
Delta R.T.: -0.018 min
Response: 992482
Conc: 0.26 ng/ml




#9 Endosulfan I

R.T.: 0.000 min
Exp R.T. : 6.070 min
Response: 0
Conc: N.D.



#9 Endosulfan I


R.T.: 5.129 min
Delta R.T.: 0.027 min
Response: 311790
Conc: 0.10 ng/ml



#10 gamma-Chlordane


R.T.: 0.000 min
Exp R.T.: 5.941 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
WC-TA2-03-C




#10 gamma-Chlordane

R.T.: 4.977 min
Delta R.T.: -0.005 min
Response: 2796866
Conc: 0.83 ng/ml




#11 alpha-Chlordane

R.T.: 0.000 min
Exp R.T.: 6.019 min
Response: 0
Conc: N.D.



#11 alpha-Chlordane


R.T.: 5.021 min
Delta R.T.: -0.025 min
Response: 4067813
Conc: 1.22 ng/ml



#13 Dieldrin


R.T.: 0.000 min
Exp R.T.: 6.345 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
WC-TA2-03-C




#13 Dieldrin

R.T.: 5.352 min
Delta R.T.: -0.014 min
Response: 1679191
Conc: 0.50 ng/ml




#14 Endrin

R.T.: 0.000 min
Exp R.T.: 6.574 min
Response: 0
Conc: N.D.



#14 Endrin


R.T.: 5.655 min
Delta R.T.: 0.014 min
Response: 1457763
Conc: 0.50 ng/ml



#16 4,4'-DDD


R.T.: 0.000 min
Exp R.T.: 6.710 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
WC-TA2-03-C




#16 4,4'-DDD

R.T.: 5.771 min
Delta R.T.: -0.018 min
Response: 548435
Conc: 0.22 ng/ml




#17 4,4'-DDT

R.T.: 0.000 min
Exp R.T.: 7.023 min
Response: 0
Conc: N.D.



#17 4,4'-DDT


R.T.: 6.030 min
Delta R.T.: -0.010 min
Response: 660611
Conc: 0.25 ng/ml



#20 Methoxychlor


R.T.: 0.000 min
Exp R.T.: 7.499 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
WC-TA2-03-C




#20 Methoxychlor

R.T.: 6.596 min
Delta R.T.: -0.019 min
Response: 449491
Conc: 0.31 ng/ml




#25 Chlordane-3

R.T.: 0.000 min
Exp R.T.: 5.941 min
Response: 0
Conc: N.D.



#25 Chlordane-3


R.T.: 4.977 min
Delta R.T.: -0.006 min
Response: 2796866
Conc: 7.71 ng/ml



#26 Chlordane-4


R.T.: 0.000 min
Exp R.T.: 6.023 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
WC-TA2-03-C




#26 Chlordane-4

R.T.: 5.021 min
Delta R.T.: -0.024 min
Response: 4067813
Conc: 11.63 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.051 min
Delta R.T.: -0.003 min
Response: 35955204
Conc: 18.68 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.913 min
Delta R.T.: -0.003 min
Response: 54137932
Conc: 19.84 ng/ml