

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL111419\
 Data File : PL054246.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Nov 2019 14:41
 Operator : SG\AJ
 Sample : K5826-07MSD
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TP-14B-SMSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 00:36:16 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 18 13:56:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.539	4.042	236.3E6	54401100	21.801	21.740
28)	SA Decachlor...	8.321	9.119	192.0E6	54542497	19.571	19.369
Target Compounds							
2)	A alpha-BHC	3.982	4.433	1073.2E6	198.9E6	61.547	57.526
3)	MA gamma-BHC...	4.269	4.719	964.9E6	187.7E6	61.959	58.300
4)	MA Heptachlor	4.575	5.228	951.6E6	170.1E6	60.293	52.800
5)	MB Aldrin	4.826	5.533	880.9E6	162.1E6	61.618	54.385
6)	B beta-BHC	4.519	4.889	390.8E6	84181468	59.299	54.500
7)	B delta-BHC	4.727	5.107	845.3E6	142.3E6	55.532	47.125
8)	B Heptachlo...	5.271	5.919	793.1E6	144.5E6	58.753	49.536
9)	A Endosulfan I	5.613	6.276	710.6E6	140.2E6	57.806	51.699
10)	B gamma-Chl...	5.500	6.153	803.3E6	143.6E6	58.879	49.226
11)	B alpha-Chl...	5.558	6.226	772.6E6	152.1E6	58.073	52.103
12)	B 4,4'-DDE	5.722	6.382	733.0E6	148.5E6	61.689	52.730
13)	MA Dieldrin	5.858	6.533	749.7E6	149.0E6	57.521	52.226
14)	MA Endrin	6.116	6.751	668.9E6	128.3E6	57.706	52.850
15)	B Endosulfa...	6.389	6.960	591.9E6	131.8E6	52.943	51.601
16)	A 4,4'-DDD	6.239	6.872	519.2E6	115.2E6	58.339	55.262
17)	MA 4,4'-DDT	6.482	7.171	551.2E6	121.2E6	55.642	52.419
18)	B Endrin al...	6.557	7.084	464.3E6	109.8E6	50.637	49.983
19)	B Endosulfa...	6.771	7.307	502.7E6	114.2E6	49.422	46.534
20)	A Methoxychlor	7.026	7.629	251.6E6	61252826	51.023	47.157
21)	B Endrin ke...	7.264	7.778	548.5E6	129.3E6	50.950	50.142
22)	Mirex	7.457	8.232	461.7E6	110.5E6	53.984	54.602

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL111419\
Data File : PL054246.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2019 14:41
Operator : SG\AJ
Sample : K5826-07MSD
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
TP-14B-SMSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 00:36:16 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
Quant Title : GC Extractables
QLast Update : Fri Oct 18 13:56:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

