

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL111422\
 Data File : PL079099.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Nov 2022 16:51
 Operator : AR\AJ
 Sample : N5607-11
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 SB-40A

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 11/15/2022
 Supervised By :Ankita Jodhani 11/15/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 14 20:23:33 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102122.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 21 07:08:41 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.321	2.568	27525142	33495652	14.434	13.119
28)	SA Decachlor...	8.769	7.634	17764071	23157909	11.808	12.226
Target Compounds							
12)	B 4,4'-DDE	5.969	4.973	11008482	16668013	5.501	6.122m
16)	A 4,4'-DDD	6.495	5.525	2898132	2794499	1.739	1.272 #
17)	MA 4,4'-DDT	6.803	5.769	4984348	8827159	2.781m	4.069m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL111422\
Data File : PL079099.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2022 16:51
Operator : AR\AJ
Sample : N5607-11
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
SB-40A

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 11/15/2022
Supervised By : Ankita Jodhani 11/15/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 20:23:33 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102122.M
Quant Title : GC Extractables
QLast Update : Fri Oct 21 07:08:41 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

