

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL111422\
 Data File : PL079098.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Nov 2022 16:37
 Operator : AR\AJ
 Sample : N5607-09
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 SB-39

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 11/15/2022
 Supervised By : Ankita Jodhani 11/15/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 14 20:23:11 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102122.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 21 07:08:41 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.320	2.590f	34376104	41766689	18.026	16.359m
28)	SA Decachlor...	8.768	7.657	23871245	28467240	15.868	15.029
Target Compounds							
12)	B 4,4'-DDE	5.967	4.995	1163084	1400424	0.581m	0.514m
16)	A 4,4'-DDD	6.493	5.546	2182706	1896318	1.309m	0.863m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL111422\
Data File : PL079098.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2022 16:37
Operator : AR\AJ
Sample : N5607-09
Misc :
ALS Vial : 16 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
SB-39


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 11/15/2022
Supervised By : Ankita Jodhani 11/15/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 20:23:11 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102122.M
Quant Title : GC Extractables
QLast Update : Fri Oct 21 07:08:41 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.320 min
Delta R.T.: -0.003 min
Response: 34376104
Conc: 18.03 ng/ml

Instrument :
ECD_L
ClientSampleId :
SB-39


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 11/15/2022
Supervised By :Ankita Jodhani 11/15/2022




#1 Tetrachloro-m-xylene

R.T.: 2.590 min
Delta R.T.: 0.016 min
Response: 41766689
Conc: 16.36 ng/ml m




#12 4,4'-DDE

R.T.: 5.967 min
Delta R.T.: -0.007 min
Response: 1163084
Conc: 0.58 ng/ml m



#12 4,4'-DDE

R.T.: 4.995 min
Delta R.T.: 0.013 min
Response: 1400424
Conc: 0.51 ng/ml m




#16 4,4'-DDD

R.T.: 6.493 min
Delta R.T.: -0.006 min
Response: 2182706
Conc: 1.31 ng/ml

Instrument :
ECD_L
ClientSampleId :
SB-39


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 11/15/2022
Supervised By :Ankita Jodhani 11/15/2022




#16 4,4'-DDD

R.T.: 5.546 min
Delta R.T.: 0.014 min
Response: 1896318
Conc: 0.86 ng/ml m



#28 Decachlorobiphenyl

R.T.: 8.768 min
Delta R.T.: -0.008 min
Response: 23871245
Conc: 15.87 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.657 min
Delta R.T.: 0.013 min
Response: 28467240
Conc: 15.03 ng/ml