

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL111424\
 Data File : PL093055.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Nov 2024 10:12
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 14 23:35:18 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102824.M
 Quant Title : GC Extractables
 QLast Update : Mon Oct 28 18:58:23 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.545	2.778	114.0E6	134.9E6	46.518	49.580
28) SA Decachlor...	9.060	7.917	83445422	119.0E6	43.358	43.597
Target Compounds						
2) A alpha-BHC	4.001	3.281	157.3E6	201.1E6	46.391	50.409
3) MA gamma-BHC...	4.334	3.611	150.1E6	193.3E6	46.064	50.092
4) MA Heptachlor	4.922	3.950	132.4E6	185.5E6	44.102	49.163
5) MB Aldrin	5.263	4.230	134.6E6	186.2E6	44.857	50.883
6) B beta-BHC	4.531	3.910	66624959	82911017	46.093	50.364
7) B delta-BHC	4.778	4.139	144.2E6	200.4E6	46.062	52.149
8) B Heptachlo...	5.689	4.732	122.4E6	168.5E6	44.321	50.439
9) A Endosulfan I	6.075	5.102	110.9E6	154.4E6	44.347	50.557
10) B gamma-Chl...	5.946	4.982	116.2E6	175.1E6	43.668	52.089
11) B alpha-Chl...	6.024	5.046	118.1E6	169.6E6	44.589	50.988
12) B 4,4'-DDE	6.198	5.235	107.6E6	167.3E6	45.425	51.936
13) MA Dieldrin	6.350	5.367	117.2E6	169.9E6	44.461	50.876
14) MA Endrin	6.580	5.641	92651884	141.2E6	40.696	48.800
15) B Endosulfa...	6.799	5.937	103.0E6	146.9E6	43.202	51.842
16) A 4,4'-DDD	6.715	5.790	92012788	136.2E6	47.935	54.685
17) MA 4,4'-DDT	7.029	6.040	86347773	127.1E6	41.719	47.400
18) B Endrin al...	6.929	6.116	83755140	117.0E6	44.599	50.710
19) B Endosulfa...	7.163	6.339	95334858	140.7E6	43.926	52.169
20) A Methoxychlor	7.505	6.616	48218980	67678119	42.297	47.393
21) B Endrin ke...	7.648	6.844	107.9E6	160.1E6	44.539	52.157
22) Mirex	8.121	7.025	84900305	130.1E6	42.842	49.881

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL111424\
Data File : PL093055.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2024 10:12
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 23:35:18 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102824.M
Quant Title : GC Extractables
QLast Update : Mon Oct 28 18:58:23 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

