

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL111519\
 Data File : PL054263.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Nov 2019 09:00
 Operator : SG\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 16 00:16:47 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 18 13:56:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.540	4.042	615.4E6	137.3E6	56.776	54.866
2)	SA Decachlor...	8.321	9.120	498.4E6	139.3E6	50.794	49.472
Target Compounds							
2)	A alpha-BHC	3.983	4.434	1019.8E6	193.1E6	58.486	55.869
3)	MA gamma-BHC...	4.270	4.719	904.7E6	177.8E6	58.093	55.198
4)	MA Heptachlor	4.576	5.229	881.0E6	160.3E6	55.822	49.764
5)	MB Aldrin	4.827	5.534	832.1E6	152.7E6	58.205	51.200
6)	B beta-BHC	4.520	4.889	369.9E6	81316251	56.128	52.645
7)	B delta-BHC	4.728	5.107	873.7E6	167.6E6	57.402	55.487
8)	B Heptachlo...	5.272	5.920	760.6E6	144.9E6	56.348	49.679
9)	A Endosulfan I	5.614	6.277	681.9E6	133.6E6	55.477	49.282
10)	B gamma-Chl...	5.500	6.154	764.6E6	146.0E6	56.041	50.035
11)	B alpha-Chl...	5.559	6.227	736.6E6	144.8E6	55.369	49.601
12)	B 4,4'-DDE	5.722	6.383	691.1E6	139.8E6	58.162	49.655
13)	MA Dieldrin	5.858	6.534	721.5E6	141.6E6	55.355	49.637
14)	MA Endrin	6.116	6.752	641.9E6	121.0E6	55.375	49.869
15)	B Endosulfa...	6.390	6.961	593.9E6	126.6E6	53.122	49.561
16)	A 4,4'-DDD	6.240	6.873	515.2E6	108.8E6	57.886	52.220
17)	MA 4,4'-DDT	6.481	7.172	518.9E6	109.7E6	52.379	47.444
18)	B Endrin al...	6.558	7.085	479.6E6	107.0E6	52.306	48.708
19)	B Endosulfa...	6.772	7.308	530.4E6	119.2E6	52.146	48.568
20)	A Methoxychlor	7.027	7.630	248.3E6	60483416	50.371	46.565
21)	B Endrin ke...	7.265	7.779	585.2E6	129.6E6	54.363	50.278
22)	Mirex	7.458	8.234	430.4E6	102.3E6	50.326	50.559

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL111519\
Data File : PL054263.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Nov 2019 09:00
Operator : SG\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 16 00:16:47 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
Quant Title : GC Extractables
QLast Update : Fri Oct 18 13:56:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

