

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL111519\
 Data File : PL054305.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Nov 2019 19:00
 Operator : SG\AJ
 Sample : K5833-03
 Misc :
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 7-(2-4)

Manual Integrations
 APPROVED

Ankita
 11/18/2019 3:16:50 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 16 00:29:07 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 18 13:56:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.539	4.043	205.6E6	48521549	18.966	19.390
28)	SA Decachlor...	8.321	9.121	160.6E6	49248328	16.370	17.489m
Target Compounds							
9)	A Endosulfan I	5.636	6.275	88624328	4129622	7.210	1.523m#
16)	A 4,4'-DDD	6.242	6.873	9107932	9285341	1.023	4.456 #
17)	MA 4,4'-DDT	6.480	7.173	27052034	4639392	2.731	2.006 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL111519\
Data File : PL054305.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Nov 2019 19:00
Operator : SG\AJ
Sample : K5833-03
Misc :
ALS Vial : 43 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampled :
7-(2-4)


Manual Integrations
APPROVED

Ankita
11/18/2019 3:16:50 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 16 00:29:07 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
Quant Title : GC Extractables
QLast Update : Fri Oct 18 13:56:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.539 min
Delta R.T.: -0.005 min
Response: 205578622
Conc: 18.97 ng/ml

Instrument :
ECD_L
ClientSampleId :
7-(2-4)


Manual Integrations
APPROVED

Ankita
11/18/2019 3:16:50 PM




#1 Tetrachloro-m-xylene

R.T.: 4.043 min
Delta R.T.: -0.003 min
Response: 48521549
Conc: 19.39 ng/ml




#9 Endosulfan I

R.T.: 5.636 min
Delta R.T.: 0.013 min
Response: 88624328
Conc: 7.21 ng/ml



#9 Endosulfan I

R.T.: 6.275 min
Delta R.T.: -0.006 min
Response: 4129622
Conc: 1.52 ng/ml m




#16 4,4'-DDD

R.T.: 6.242 min
Delta R.T.: -0.007 min
Response: 9107932
Conc: 1.02 ng/ml

Instrument :
ECD_L
ClientSampleId :
7-(2-4)


Manual Integrations
APPROVED

Ankita
11/18/2019 3:16:50 PM




#16 4,4'-DDD

R.T.: 6.873 min
Delta R.T.: -0.003 min
Response: 9285341
Conc: 4.46 ng/ml




#17 4,4'-DDT

R.T.: 6.480 min
Delta R.T.: -0.011 min
Response: 27052034
Conc: 2.73 ng/ml



#17 4,4'-DDT

R.T.: 7.173 min
Delta R.T.: -0.002 min
Response: 4639392
Conc: 2.01 ng/ml




#28 Decachlorobiphenyl

R.T.: 8.321 min
Delta R.T.: -0.013 min
Response: 160611932
Conc: 16.37 ng/ml

Instrument :
ECD_L
ClientSampleId :
7-(2-4)

Manual Integrations
APPROVED

Ankita
11/18/2019 3:16:50 PM



#28 Decachlorobiphenyl

R.T.: 9.121 min
Delta R.T.: -0.005 min
Response: 49248328
Conc: 17.49 ng/ml m