

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL111618\
 Data File : PL042209.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Nov 2018 20:49
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 11/15/2018 11:54:00 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 05:45:49 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL110918.M
 Quant Title : GC Extractables
 QLast Update : Fri Nov 09 12:38:40 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.468	4.101	269.5E6	135.2E6	43.648	45.620
28)	SA Decachlor...	8.203	9.228	274.3E6	112.0E6	40.045m	43.910
Target Compounds							
2)	A alpha-BHC	3.907	4.496	425.1E6	202.8E6	44.836	45.686
3)	MA gamma-BHC...	4.190	4.784	384.2E6	185.1E6	44.527	44.900
4)	MA Heptachlor	4.486	5.300	401.2E6	181.5E6	45.064	45.083
5)	MB Aldrin	4.732	5.608	357.5E6	170.9E6	43.208	45.462
6)	B beta-BHC	4.444	4.953	160.9E6	76179254	43.603	46.353
7)	B delta-BHC	4.645	5.172	361.9E6	181.2E6	44.697	45.833m
8)	B Heptachlo...	5.175	5.996	338.5E6	157.0E6	42.910	43.702
9)	A Endosulfan I	5.513	6.356	302.2E6	143.6E6	41.472	44.019
10)	B gamma-Chl...	5.401	6.231	344.6E6	156.4E6	42.757	44.130
11)	B alpha-Chl...	5.458	6.305	331.2E6	152.0E6	41.561	43.693
12)	B 4,4'-DDE	5.624	6.459	284.9E6	152.4E6	41.641	44.636
13)	MA Dieldrin	5.755	6.614	332.0E6	151.7E6	42.676	45.314
14)	MA Endrin	6.011	6.834	306.7E6	129.4E6	43.118	43.973
15)	B Endosulfa...	6.286	7.042	298.6E6	130.1E6	43.747	43.774
16)	A 4,4'-DDD	6.140	6.950	244.1E6	118.1E6	44.162	44.789
17)	MA 4,4'-DDT	6.378	7.251	263.4E6	120.9E6	46.304	44.854
18)	B Endrin al...	6.454	7.166	244.1E6	108.3E6	43.654	43.698
19)	B Endosulfa...	6.666	7.390	281.4E6	120.2E6	44.130	44.180
20)	A Methoxychlor	6.924	7.707	147.0E6	62731504	46.142	45.653
21)	B Endrin ke...	7.156	7.864	318.2E6	127.1E6	44.153	43.898

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL111618\
Data File : PL042209.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2018 20:49
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
11/15/2018 11:54:00 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 05:45:49 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL110918.M
Quant Title : GC Extractables
QLast Update : Fri Nov 09 12:38:40 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

