

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL111718\
 Data File : PL042238.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Nov 2018 12:24
 Operator : AJ\SJ
 Sample : MIREX-MDL-S-01
 Misc : FOR MIREX
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 MIREX-MDL-S-01

Manual Integrations
 APPROVED

Sohil
 11/16/2018 4:01:36 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 13:11:22 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\Mirex-PL111718.M
 Quant Title : GC Extractables
 QLast Update : Thu Nov 15 11:21:10 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.468	4.101	98049570	51143775	17.665	18.501
28)	SA Decachlor...	8.205	9.230	119.3E6	46076568	19.281	19.356
Target Compounds							
22)	Mirex	7.338	8.321	6476190	1642305	1.136	0.881m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL111718\
Data File : PL042238.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Nov 2018 12:24
Operator : AJ\SJ
Sample : MIREX-MDL-S-01
Misc : FOR MIREX
ALS Vial : 12 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
MIREX-MDL-S-01


Manual Integrations
APPROVED

Sohil
11/16/2018 4:01:36 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 13:11:22 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\Mirex-PL111718.M
Quant Title : GC Extractables
QLast Update : Thu Nov 15 11:21:10 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene

R.T.: 3.468 min
Delta R.T.: 0.000 min
Response: 98049570
Conc: 17.67 ng/ml

Instrument :
ECD_L
ClientSampleId :
MIREX-MDL-S-01


Manual Integrations
APPROVED

Sohil
11/16/2018 4:01:36 PM




#1 Tetrachloro-m-xylene

R.T.: 4.101 min
Delta R.T.: 0.000 min
Response: 51143775
Conc: 18.50 ng/ml




#22 Mirex

R.T.: 7.338 min
Delta R.T.: 0.001 min
Response: 6476190
Conc: 1.14 ng/ml



#22 Mirex

R.T.: 8.321 min
Delta R.T.: 0.000 min
Response: 1642305
Conc: 0.88 ng/ml m




#28 Decachlorobiphenyl

R.T.: 8.205 min
Delta R.T.: 0.002 min
Response: 119317313
Conc: 19.28 ng/ml

Instrument :
ECD_L
ClientSampleId :
MIREX-MDL-S-01

Manual Integrations
APPROVED

Sohil
11/16/2018 4:01:36 PM



#28 Decachlorobiphenyl

R.T.: 9.230 min
Delta R.T.: 0.001 min
Response: 46076568
Conc: 19.36 ng/ml