

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL112118\
 Data File : PL042358.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Nov 2018 17:00
 Operator : AJ\SJ
 Sample : PB114977BL
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 21 00:04:57 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112018.M
 Quant Title : GC Extractables
 QLast Update : Tue Nov 20 02:29:59 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.468	4.102	108.2E6	58267640	18.388	20.259
28)	SA Decachlor...	8.206	9.232	121.9E6	48297960	19.350	20.025
Target Compounds							
2)	A alpha-BHC	3.902	4.504	3048976	1421186	0.338	0.332
6)	B beta-BHC	4.403f	0.000	4035463	0	1.139	N.D. #
8)	B Heptachlo...	0.000	5.996	0	2119929	N.D.	0.591 #
19)	B Endosulfa...	6.656	7.385	4345464	2975798	0.696	1.070 #
20)	A Methoxychlor	0.000	7.648f	0	4483759	N.D.	3.289 #
23)	Chlordane-1	4.364	0.000	1321850	0	4.781	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL112118\
Data File : PL042358.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Nov 2018 17:00
Operator : AJ\SJ
Sample : PB114977BL
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 00:04:57 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112018.M
Quant Title : GC Extractables
QLast Update : Tue Nov 20 02:29:59 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.468 min
Delta R.T.: 0.000 min
Response: 108187213
Conc: 18.39 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.102 min
Delta R.T.: 0.000 min
Response: 58267640
Conc: 20.26 ng/ml




#2 alpha-BHC

R.T.: 3.902 min
Delta R.T.: -0.006 min
Response: 3048976
Conc: 0.34 ng/ml



#2 alpha-BHC


R.T.: 4.504 min
Delta R.T.: 0.007 min
Response: 1421186
Conc: 0.33 ng/ml



#6 beta-BHC


R.T.: 4.403 min
Delta R.T.: -0.041 min
Response: 4035463
Conc: 1.14 ng/ml

Instrument :
ECD_L
ClientSampleId :




#6 beta-BHC

R.T.: 0.000 min
Exp R.T. : 4.954 min
Response: 0
Conc: N.D.




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T. : 5.177 min
Response: 0
Conc: N.D.



#8 Heptachlor epoxide


R.T.: 5.996 min
Delta R.T.: -0.002 min
Response: 2119929
Conc: 0.59 ng/ml



#19 Endosulfan Sulfate


R.T.: 6.656 min
Delta R.T.: -0.013 min
Response: 4345464
Conc: 0.70 ng/ml

Instrument :
ECD_L
ClientSampleId :




#19 Endosulfan Sulfate

R.T.: 7.385 min
Delta R.T.: -0.008 min
Response: 2975798
Conc: 1.07 ng/ml




#20 Methoxychlor

R.T.: 0.000 min
Exp R.T. : 6.926 min
Response: 0
Conc: N.D.



#20 Methoxychlor


R.T.: 7.648 min
Delta R.T.: -0.061 min
Response: 4483759
Conc: 3.29 ng/ml



#23 Chlordane-1


R.T.: 4.364 min
Delta R.T.: 0.029 min
Response: 1321850
Conc: 4.78 ng/ml

Instrument :
ECD_L
ClientSampleId :




#23 Chlordane-1

R.T.: 0.000 min
Exp R.T. : 5.111 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 8.206 min
Delta R.T.: 0.000 min
Response: 121947800
Conc: 19.35 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.232 min
Delta R.T.: 0.000 min
Response: 48297960
Conc: 20.03 ng/ml