

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL112218\
 Data File : PL042401.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Nov 2018 08:01
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 22 00:40:26 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112018.M
 Quant Title : GC Extractables
 QLast Update : Tue Nov 20 02:29:59 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.467	4.101	279.0E6	142.4E6	47.416	49.512
28)	SA Decachlor...	8.205	9.231	271.3E6	106.3E6	43.055	44.086
Target Compounds							
2)	A alpha-BHC	3.907	4.496	434.3E6	213.8E6	48.130	49.921
3)	MA gamma-BHC...	4.190	4.785	396.2E6	196.1E6	48.070	49.616
4)	MA Heptachlor	4.486	5.300	394.2E6	183.8E6	45.306	46.847
5)	MB Aldrin	4.732	5.609	386.9E6	185.3E6	47.659	48.240
6)	B beta-BHC	4.444	4.954	165.6E6	80657742	46.745	51.123
7)	B delta-BHC	4.646	5.174	373.0E6	184.1E6	48.706	50.020
8)	B Heptachlo...	5.175	5.998	355.1E6	166.2E6	46.848	46.356
9)	A Endosulfan I	5.512	6.358	323.9E6	151.0E6	46.799	47.809
10)	B gamma-Chl...	5.400	6.232	357.5E6	163.9E6	46.893	48.139
11)	B alpha-Chl...	5.458	6.306	349.7E6	159.2E6	46.591	47.617
12)	B 4,4'-DDE	5.624	6.460	312.6E6	159.9E6	47.761	48.033
13)	MA Dieldrin	5.755	6.616	348.5E6	158.0E6	47.375	47.465
14)	MA Endrin	6.011	6.835	262.8E6	110.3E6	40.177	40.028
15)	B Endosulfa...	6.286	7.044	296.2E6	130.3E6	47.381	47.192
16)	A 4,4'-DDD	6.140	6.953	265.3E6	127.0E6	50.101	48.233
17)	MA 4,4'-DDT	6.378	7.253	251.6E6	112.8E6	45.194	42.730
18)	B Endrin al...	6.454	7.168	247.6E6	109.8E6	47.622	47.910
19)	B Endosulfa...	6.667	7.391	295.6E6	127.0E6	47.384	45.652
20)	A Methoxychlor	6.925	7.709	133.8E6	55638161	42.416	40.807
21)	B Endrin ke...	7.157	7.866	337.1E6	134.6E6	48.686	48.691

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL112218\
Data File : PL042401.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Nov 2018 08:01
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 22 00:40:26 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112018.M
Quant Title : GC Extractables
QLast Update : Tue Nov 20 02:29:59 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

