

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : DTX112322.M
 Title : GC Extractables
 Last Update : Thu Nov 24 00:56:34 2022
 Response Via : Initial Calibration

Calibration Files

500 =PL079406.D 1000 =PL079404.D 750 =PL079405.D
 250 =PL079407.D 100 =PL079408.D

	Compound	500	1000	750	250	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	2.257	2.300	2.235	2.184	2.153	2.226	E6	2.62
2)	Toxaphene-1	1.514	1.637	1.613	1.420	1.369	1.510	E4	7.76
3)	Toxaphene-2	6.875	6.748	6.532	6.512	6.604	6.654	E4	2.32
4)	Toxaphene-3	5.357	5.347	5.212	4.991	4.719	5.125	E4	5.29
5)	Toxaphene-4	3.755	3.833	3.696	3.561	3.453	3.660	E4	4.17
6)	Toxaphene-5	3.291	3.600	3.365	3.453	3.123	3.367	E4	5.30
7) SA	Decachlorobiphenyl	1.925	1.891	1.857	1.928	1.957	1.912	E6	2.02

Signal #2 Calibration Files

500 =PL079406.D 1000 =PL079404.D 750 =PL079405.D
 250 =PL079407.D 100 =PL079408.D

	Compound	500	1000	750	250	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	2.711	2.700	2.651	2.697	2.700	2.692	E6	0.87
2)	Toxaphene-1	1.652	1.652	1.615	1.640	1.485	1.609	E4	4.41
3)	Toxaphene-2	2.434	2.305	2.263	2.320	2.314	2.327	E4	2.73
4)	Toxaphene-3	6.874	7.005	6.792	6.530	6.329	6.706	E4	4.08
5)	Toxaphene-4	1.003	0.993	0.958	0.985	0.927	0.973	E5	3.16
6)	Toxaphene-5	7.052	6.875	6.840	6.780	6.508	6.811	E4	2.89
7) SA	Decachlorobiphenyl	2.289	2.224	2.187	2.341	2.456	2.299	E6	4.60

(#) = Out of Range