

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL112322\
 Data File : PL079404.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Nov 2022 21:18
 Operator : AR\AJ
 Sample : PTOXICC1000
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PTOXICC1000

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 11/25/2022
 Supervised By :Ankita Jodhani 11/25/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 24 00:54:39 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX112322.M
 Quant Title : GC Extractables
 QLast Update : Thu Nov 24 00:54:30 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds						
1) SA Tetrachlo...	3.318	2.564	230.0E6	270.0E6	103.322	100.314
7) SA Decachlor...	8.765	7.630	189.1E6	222.4E6	98.920	96.726
Target Compounds						
2) Toxaphene-1	6.006	5.060	16372659	16520269	1013.380	1026.922
3) Toxaphene-2	6.829	5.693	67478312	23048424	1016.044m	990.316
4) Toxaphene-3	6.919	6.318	53474340	70054953	1043.344	1044.673
5) Toxaphene-4	7.009	6.445	38331161	99291459	1047.414	1020.308
6) Toxaphene-5	7.346	6.754	36004916	68745274	1069.484	1009.337

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL112322\
Data File : PL079404.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Nov 2022 21:18
Operator : AR\AJ
Sample : PTOXICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 24 00:54:39 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX112322.M
Quant Title : GC Extractables
QLast Update : Thu Nov 24 00:54:30 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Reviewed By: Abdul Mirza 11/25/2022
Supervised By: Ankita Jodhani 11/25/2022

Manual Integrations
APPROVED

Instrument :
ECD_L
Client Sampled :
PTOXICC1000