

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL112524\
 Data File : PL093250.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Nov 2024 15:16
 Operator : AR\AJ
 Sample : PTOXICV500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 ICVPL112524

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 25 15:25:37 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\LTX112524.M
 Quant Title : GC Extractables
 QLast Update : Mon Nov 25 14:58:41 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticide 1 Signal #2 Phase: Rtx-CLPesticide 1
 Signal #1 Info : 30M x 0.32mm x0.3 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.536	2.773	131.0E6	148.7E6	49.264	49.881
7) SA Decachlor...	9.054	7.912	93410824	152.0E6	51.011	50.082
Target Compounds						
2) Toxaphene-1	6.234	5.001	12144120	11195448	485.466	493.902
3) Toxaphene-2	6.438	5.326	7342033	11738327	493.629	522.596
4) Toxaphene-3	7.056	5.685	38962175	12257339	493.622	511.999
5) Toxaphene-4	7.148	6.600	29268978	38486297	497.217	518.448
6) Toxaphene-5	7.932	7.041	21249303	37044403	491.440	501.727

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL112524\
Data File : PL093250.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Nov 2024 15:16
Operator : AR\AJ
Sample : PTOXICV500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
ICVPL112524

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 25 15:25:37 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\LTX112524.M
Quant Title : GC Extractables
QLast Update : Mon Nov 25 14:58:41 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticide 1 Signal #2 Phase: Rtx-CLPesticide 1
Signal #1 Info : 30M x 0.32mm x 0.3 Signal #2 Info : 30M x 0.32mm x 0.25µm

