

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL112524\
 Data File : PL093253.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Nov 2024 15:56
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 25 17:20:11 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112524.M
 Quant Title : GC Extractables
 QLast Update : Mon Nov 25 15:18:43 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.536	2.773	132.2E6	150.0E6	50.861	51.999
28)	SA Decachlor...	9.055	7.912	91807685	152.2E6	52.807	53.272
Target Compounds							
2)	A alpha-BHC	3.992	3.276	181.8E6	225.2E6	50.949	52.758
3)	MA gamma-BHC...	4.325	3.606	171.8E6	218.1E6	50.857	52.692
4)	MA Heptachlor	4.913	3.945	154.1E6	212.9E6	50.411	52.630
5)	MB Aldrin	5.255	4.224	149.2E6	209.9E6	49.624	52.757
6)	B beta-BHC	4.522	3.906	75363666	90971433	49.919	51.140
7)	B delta-BHC	4.770	4.135	164.2E6	220.3E6	49.625	51.625
8)	B Heptachlo...	5.681	4.727	136.0E6	190.9E6	48.995	52.411
9)	A Endosulfan I	6.067	5.097	122.3E6	172.4E6	50.231	51.559
10)	B gamma-Chl...	5.937	4.977	128.3E6	195.0E6	49.785	52.627
11)	B alpha-Chl...	6.016	5.041	129.8E6	190.1E6	50.081	52.364
12)	B 4,4'-DDE	6.190	5.230	118.5E6	183.3E6	50.672	51.195
13)	MA Dieldrin	6.342	5.361	130.3E6	189.2E6	50.817	51.335
14)	MA Endrin	6.572	5.637	107.0E6	161.6E6	51.027	50.663
15)	B Endosulfa...	6.792	5.932	111.8E6	163.1E6	51.269	51.468
16)	A 4,4'-DDD	6.708	5.785	94399403	145.2E6	51.520	51.802
17)	MA 4,4'-DDT	7.022	6.035	100.0E6	153.2E6	51.884	51.712
18)	B Endrin al...	6.922	6.111	92229349	134.9E6	51.051	51.426
19)	B Endosulfa...	7.157	6.334	104.7E6	153.6E6	50.527	50.536
20)	A Methoxychlor	7.498	6.611	54493907	78811802	52.153	51.614
21)	B Endrin ke...	7.642	6.839	118.5E6	182.6E6	52.233	54.394
22)	Mirex	8.115	7.020	93385584	142.0E6	51.702	52.857

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL112524\
Data File : PL093253.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Nov 2024 15:56
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 25 17:20:11 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112524.M
Quant Title : GC Extractables
QLast Update : Mon Nov 25 15:18:43 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

