

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL112525\
 Data File : PL097753.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Nov 2025 16:27
 Operator : AR\AJ
 Sample : PSTDICV050
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 ICVPL112525

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 25 18:01:52 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112525.M
 Quant Title : GC Extractables
 QLast Update : Tue Nov 25 17:56:07 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.504	2.802	261.9E6	362.5E6	51.108	50.678
28)	SA Decachlor...	8.969	7.951	159.8E6	240.2E6	49.785	51.034
Target Compounds							
2)	A alpha-BHC	3.951	3.305	447.4E6	599.3E6	52.734	52.317
3)	MA gamma-BHC...	4.279	3.636	408.1E6	539.0E6	52.351	52.508
4)	MA Heptachlor	4.869	3.981	387.7E6	496.5E6	51.974	52.473
5)	MB Aldrin	5.208	4.262	357.6E6	493.1E6	52.116	51.979
6)	B beta-BHC	4.469	3.932	167.1E6	226.3E6	51.181	51.155
7)	B delta-BHC	4.713	4.163	373.6E6	535.2E6	53.533	52.234
8)	B Heptachlo...	5.628	4.763	309.4E6	433.3E6	49.193	51.627
9)	A Endosulfan I	6.009	5.134	275.9E6	425.5E6	51.142	49.197
10)	B gamma-Chl...	5.882	5.015	306.9E6	467.2E6	51.354	52.201
11)	B alpha-Chl...	5.961	5.079	301.7E6	461.4E6	50.435	49.280
12)	B 4,4'-DDE	6.133	5.267	274.9E6	422.2E6	51.308	51.873
13)	MA Dieldrin	6.281	5.398	288.0E6	443.7E6	51.493	51.348
14)	MA Endrin	6.508	5.671	233.3E6	383.8E6	50.687	49.939
15)	B Endosulfa...	6.722	5.964	232.4E6	375.6E6	45.816	51.592
16)	A 4,4'-DDD	6.642	5.819	214.8E6	354.3E6	50.450	52.375
17)	MA 4,4'-DDT	6.956	6.071	199.0E6	291.2E6	54.545	59.087
18)	B Endrin al...	6.850	6.142	178.5E6	294.9E6	49.144	54.160
19)	B Endosulfa...	7.084	6.365	220.7E6	354.7E6	50.164	51.690
20)	A Methoxychlor	7.429	6.643	100.3E6	148.9E6	53.150	55.427
21)	B Endrin ke...	7.562	6.868	226.0E6	380.9E6	51.779	52.282
22)	Mirex	8.038	7.056	153.8E6	263.3E6	50.614	50.891

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL112525\
Data File : PL097753.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Nov 2025 16:27
Operator : AR\AJ
Sample : PSTDICV050
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
ICVPL112525

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 25 18:01:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112525.M
Quant Title : GC Extractables
QLast Update : Tue Nov 25 17:56:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

