

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL112618\
 Data File : PL042459.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Nov 2018 16:21
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 27 00:33:07 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112018.M
 Quant Title : GC Extractables
 QLast Update : Tue Nov 20 02:29:59 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.467	4.102	308.0E6	159.9E6	52.344	55.605
28)	SA Decachlor...	8.208	9.236	285.3E6	121.7E6	45.268	50.452
Target Compounds							
2)	A alpha-BHC	3.907	4.497	485.2E6	241.8E6	53.773	56.457
3)	MA gamma-BHC...	4.191	4.786	440.5E6	221.2E6	53.449	55.966
4)	MA Heptachlor	4.486	5.302	442.4E6	213.8E6	50.845	54.500
5)	MB Aldrin	4.733	5.611	429.7E6	213.7E6	52.926	55.650
6)	B beta-BHC	4.444	4.955	182.2E6	88543888	51.429	56.121
7)	B delta-BHC	4.647	5.176	415.8E6	207.7E6	54.302	56.439
8)	B Heptachlor...	5.176	6.000	390.4E6	192.5E6	51.508	53.685
9)	A Endosulfan I	5.514	6.360	351.2E6	173.9E6	50.745	55.054
10)	B gamma-Chl...	5.402	6.235	394.7E6	189.2E6	51.766	55.582
11)	B alpha-Chl...	5.459	6.308	380.4E6	183.4E6	50.672	54.852
12)	B 4,4'-DDE	5.626	6.463	340.1E6	184.6E6	51.964	55.471
13)	MA Dieldrin	5.757	6.619	377.9E6	185.8E6	51.373	55.827
14)	MA Endrin	6.013	6.838	335.6E6	151.5E6	51.312	54.983
15)	B Endosulfa...	6.289	7.046	317.8E6	152.0E6	50.849	55.031
16)	A 4,4'-DDD	6.142	6.955	281.7E6	146.2E6	53.210	55.529
17)	MA 4,4'-DDT	6.380	7.256	287.4E6	143.9E6	51.628	54.511
18)	B Endrin al...	6.457	7.171	257.7E6	124.0E6	49.562	54.087
19)	B Endosulfa...	6.669	7.394	318.0E6	148.7E6	50.972	53.470
20)	A Methoxychlor	6.927	7.712	154.6E6	71280995	49.034	52.280
21)	B Endrin ke...	7.159	7.870	348.7E6	147.7E6	50.352	53.449

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL112618\
Data File : PL042459.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Nov 2018 16:21
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 27 00:33:07 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112018.M
Quant Title : GC Extractables
QLast Update : Tue Nov 20 02:29:59 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

