

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL112619\
 Data File : PL054624.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Nov 2019 13:02
 Operator : SG\AJ
 Sample : PSTDICV050
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 ICVPL112619

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 27 02:20:59 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112619.M
 Quant Title : GC Extractables
 QLast Update : Wed Nov 27 01:19:55 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.537	4.051	554.2E6	118.7E6	47.638	46.506
28)	SA Decachlor...	8.319	9.151	450.4E6	110.8E6	47.030	46.220
Target Compounds							
2)	A alpha-BHC	3.981	4.443	910.9E6	167.4E6	47.982	47.248
3)	MA gamma-BHC...	4.267	4.728	826.7E6	152.7E6	48.499	46.824
4)	MA Heptachlor	4.573	5.238	839.9E6	139.4E6	47.629	46.182
5)	MB Aldrin	4.824	5.543	786.4E6	129.1E6	48.244	46.347
6)	B beta-BHC	4.518	4.899	346.6E6	65562187	46.732	46.557
7)	B delta-BHC	4.726	5.117	785.0E6	130.5E6	48.294	48.961
8)	B Heptachlo...	5.270	5.930	726.6E6	117.8E6	48.036	47.155
9)	A Endosulfan I	5.611	6.288	652.5E6	114.4E6	47.852	46.250
10)	B gamma-Chl...	5.498	6.164	725.1E6	122.2E6	47.820	46.394
11)	B alpha-Chl...	5.557	6.237	702.5E6	121.5E6	47.640	46.275
12)	B 4,4'-DDE	5.720	6.393	639.3E6	114.5E6	48.048	46.236
13)	MA Dieldrin	5.856	6.545	693.4E6	118.0E6	48.900	46.365
14)	MA Endrin	6.115	6.764	612.6E6	102.6E6	48.106	46.330
15)	B Endosulfa...	6.388	6.973	577.0E6	108.7E6	47.689	46.241
16)	A 4,4'-DDD	6.237	6.884	467.3E6	88735129	47.848	46.215
17)	MA 4,4'-DDT	6.480	7.184	511.2E6	92056559	48.799	46.462
18)	B Endrin al...	6.556	7.098	469.8E6	91046915	47.438	46.448
19)	B Endosulfa...	6.769	7.321	512.4E6	98466926	47.845	46.436
20)	A Methoxychlor	7.025	7.645	240.6E6	52404373	48.010	45.941
21)	B Endrin ke...	7.263	7.796	528.7E6	105.8E6	48.644	47.015
22)	Mirex	7.456	8.254	380.9E6	80271291	46.921	46.445

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL112619\
Data File : PL054624.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Nov 2019 13:02
Operator : SG\AJ
Sample : PSTDICV050
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
ICVPL112619

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 27 02:20:59 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112619.M
Quant Title : GC Extractables
QLast Update : Wed Nov 27 01:19:55 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

