

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL113020\
 Data File : PL063276.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Nov 2020 16:01
 Operator : DD\AJ
 Sample : L4891-04
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PAV-B3-112420-0-10

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 01 01:17:56 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112320.M
 Quant Title : GC Extractables
 QLast Update : Tue Nov 24 01:53:58 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.411	3.951	36646546	71629948	21.518	21.947
28)	SA Decachlor...	8.130	8.996	37379104	47184016	20.016	19.832

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL113020\
Data File : PL063276.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Nov 2020 16:01
Operator : DD\AJ
Sample : L4891-04
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PAV-B3-112420-0-10

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 01 01:17:56 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112320.M
Quant Title : GC Extractables
QLast Update : Tue Nov 24 01:53:58 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.411 min
Delta R.T.: 0.000 min
Response: 36646546
Conc: 21.52 ng/ml

Instrument :
ECD_L
ClientSampleId :
PAV-B3-112420-0-10




#1 Tetrachloro-m-xylene

R.T.: 3.951 min
Delta R.T.: 0.001 min
Response: 71629948
Conc: 21.95 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.130 min
Delta R.T.: 0.004 min
Response: 37379104
Conc: 20.02 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.996 min
Delta R.T.: 0.006 min
Response: 47184016
Conc: 19.83 ng/ml