

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL120318\
 Data File : PL042724.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Dec 2018 00:05
 Operator : AJ\SJ
 Sample : PB115250TB
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 04 02:25:08 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112018.M
 Quant Title : GC Extractables
 QLast Update : Tue Nov 20 02:29:59 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.467	4.102	113.8E6	61409063	19.342	21.351
28)	SA Decachlor...	8.206	9.233	120.0E6	51350366	19.049	21.291
Target Compounds							
2)	A alpha-BHC	3.900	0.000	3178799	0	0.352	N.D. #
6)	B beta-BHC	4.409f	4.904f	2832792	243042	0.800	0.154 #
7)	B delta-BHC	0.000	5.162	0	2335663	N.D.	0.635 #
8)	B Heptachlo...	0.000	5.995	0	5035446	N.D.	1.404 #
23)	Chlordane-1	0.000	5.162	0	2335663	N.D.	17.570 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL120318\
Data File : PL042724.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Dec 2018 00:05
Operator : AJ\SJ
Sample : PB115250TB
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 04 02:25:08 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112018.M
Quant Title : GC Extractables
QLast Update : Tue Nov 20 02:29:59 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.467 min
Delta R.T.: 0.000 min
Response: 113797811
Conc: 19.34 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.102 min
Delta R.T.: 0.000 min
Response: 61409063
Conc: 21.35 ng/ml




#2 alpha-BHC

R.T.: 3.900 min
Delta R.T.: -0.008 min
Response: 3178799
Conc: 0.35 ng/ml



#2 alpha-BHC


R.T.: 0.000 min
Exp R.T. : 4.497 min
Response: 0
Conc: N.D.



#6 beta-BHC


R.T.: 4.409 min
Delta R.T.: -0.036 min
Response: 2832792
Conc: 0.80 ng/ml

Instrument :
ECD_L
ClientSampleId :




#6 beta-BHC

R.T.: 4.904 min
Delta R.T.: -0.050 min
Response: 243042
Conc: 0.15 ng/ml




#7 delta-BHC

R.T.: 0.000 min
Exp R.T. : 4.646 min
Response: 0
Conc: N.D.



#7 delta-BHC


R.T.: 5.162 min
Delta R.T.: -0.013 min
Response: 2335663
Conc: 0.63 ng/ml



#8 Heptachlor epoxide


R.T.: 0.000 min
Exp R.T. : 5.177 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#8 Heptachlor epoxide

R.T.: 5.995 min
Delta R.T.: -0.003 min
Response: 5035446
Conc: 1.40 ng/ml




#23 Chlordane-1

R.T.: 0.000 min
Exp R.T. : 4.335 min
Response: 0
Conc: N.D.



#23 Chlordane-1


R.T.: 5.162 min
Delta R.T.: 0.051 min
Response: 2335663
Conc: 17.57 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.206 min
Delta R.T.: -0.001 min
Response: 120048952
Conc: 19.05 ng/ml

Instrument :
ECD_L
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 9.233 min
Delta R.T.: 0.002 min
Response: 51350366
Conc: 21.29 ng/ml