

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL120418\
 Data File : PL042806.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Dec 2018 03:49
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampled :
 PSTDCCC050

Manual Integrations
 APPROVED

mohammad
 12/5/2018 4:44:06 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 05 04:14:23 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112018.M
 Quant Title : GC Extractables
 QLast Update : Tue Nov 20 02:29:59 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.467	4.102	274.6E6	142.6E6	46.673	49.568
28)	SA Decachlor...	8.203	9.232	280.8E6	115.3E6	44.558	47.795
Target Compounds							
2)	A alpha-BHC	3.906	4.497	432.9E6	216.8E6	47.980	50.617
3)	MA gamma-BHC...	4.189	4.786	391.8E6	198.1E6	47.542	50.129
4)	MA Heptachlor	4.485	5.302	372.9E6	179.9E6	42.855	45.858
5)	MB Aldrin	4.730	5.610	385.4E6	191.6E6	47.466	49.894
6)	B beta-BHC	4.443	4.955	163.5E6	81204935	46.149	51.470
7)	B delta-BHC	4.645	5.176	370.0E6	189.8E6	48.318	51.587
8)	B Heptachlor...	5.174	5.998	355.4E6	173.1E6	46.896	48.277
9)	A Endosulfan I	5.511	6.359	326.7E6	159.1E6	47.209	50.369
10)	B gamma-Chl...	5.399	6.233	359.7E6	170.8E6	47.175	50.174
11)	B alpha-Chl...	5.457	6.307	352.6E6	166.8E6	46.974	49.896
12)	B 4,4'-DDE	5.623	6.461	317.5E6	168.5E6	48.510	50.642
13)	MA Dieldrin	5.754	6.617	350.5E6	170.6E6	47.646	51.252
14)	MA Endrin	6.009	6.836	263.3E6	117.6E6	40.256m	42.694
15)	B Endosulfa...	6.285	7.045	298.0E6	142.2E6	47.673	51.497
16)	A 4,4'-DDD	6.139	6.953	275.8E6	141.4E6	52.091	53.721
17)	MA 4,4'-DDT	6.375	7.254	222.9E6	113.4E6	40.032m	42.937
18)	B Endrin al...	6.452	7.169	253.7E6	121.2E6	48.805	52.879
19)	B Endosulfa...	6.665	7.393	300.2E6	139.5E6	48.107	50.154
20)	A Methoxychlor	6.921	7.710	127.5E6	57076480	40.412m	41.862
21)	B Endrin ke...	7.155	7.867	353.2E6	151.1E6	51.010	54.669

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL120418\
Data File : PL042806.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Dec 2018 03:49
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

mohammad
12/5/2018 4:44:06 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 05 04:14:23 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112018.M
Quant Title : GC Extractables
QLast Update : Tue Nov 20 02:29:59 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

