

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL120423\
 Data File : PL086966.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Dec 2023 19:24
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/05/2023
 Supervised By :Ankita Jodhani 12/05/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 04 20:39:19 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112923.M
 Quant Title : GC Extractables
 QLast Update : Wed Nov 29 06:35:04 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.342	2.609	158.4E6	44310741	49.555	48.721
28)	SA Decachlor...	8.843	7.761	133.5E6	53746670	50.012	48.747
Target Compounds							
2)	A alpha-BHC	3.796	3.108	234.2E6	63330630	50.593	49.869
3)	MA gamma-BHC...	4.128	3.435	220.8E6	61131209	50.473	50.032
4)	MA Heptachlor	4.715	3.770	216.9E6	61810613	54.651	52.936
5)	MB Aldrin	5.055	4.046	213.2E6	60213947	50.514	49.334
6)	B beta-BHC	4.332	3.738	93043916	26582739	49.739	50.243
7)	B delta-BHC	4.575	3.962	212.1E6	59104142	49.970	50.612
8)	B Heptachlo...	5.485	4.551	187.9E6	55745436	53.033	48.969
9)	A Endosulfan I	5.869	4.919	178.0E6	53909231	49.418	41.708
10)	B gamma-Chl...	5.743	4.804	192.7E6	56623518	51.240	49.397
11)	B alpha-Chl...	5.822	4.868	191.3E6	57723623	50.807	43.408
12)	B 4,4'-DDE	6.006	5.067	158.0E6	47432063	52.267	51.665
13)	MA Dieldrin	6.146	5.186	181.3E6	54016417	52.216	50.438
14)	MA Endrin	6.375	5.459	147.0E6	46764664	57.016	55.487
15)	B Endosulfa...	6.598	5.758	150.7E6	47926763	52.800	51.615
16)	A 4,4'-DDD	6.526	5.624	126.3E6	40052834	53.939	51.860
17)	MA 4,4'-DDT	6.841	5.874	132.5E6	42571651	51.627	52.788
18)	B Endrin al...	6.729	5.939	111.2E6	37492700	52.135m	53.667
19)	B Endosulfa...	6.966	6.164	148.5E6	48320620	52.007	50.927
20)	A Methoxychlor	7.328	6.458	68608251	25927862	51.957	50.373
21)	B Endrin ke...	7.448	6.665	153.6E6	57459050	51.384	50.337
22)	Mirex	7.920	6.842	128.6E6	50777592	50.265	48.925

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL120423\
Data File : PL086966.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Dec 2023 19:24
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 12/05/2023
Supervised By :Ankita Jodhani 12/05/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 04 20:39:19 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112923.M
Quant Title : GC Extractables
QLast Update : Wed Nov 29 06:35:04 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

