

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\pl120522\
 Data File : PL079559.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Dec 2022 10:31
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 06 01:14:28 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112322.M
 Quant Title : GC Extractables
 QLast Update : Thu Nov 24 00:09:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.325	2.567	115.3E6	137.3E6	51.026	50.431
2)	SA Decachlor...	8.774	7.637	99796708	107.3E6	51.386	45.672
Target Compounds							
2)	A alpha-BHC	3.786	3.054	183.4E6	221.3E6	53.347	51.332
3)	MA gamma-BHC...	4.119	3.375	173.4E6	201.0E6	53.515	51.619
4)	MA Heptachlor	4.689	3.702	168.0E6	195.3E6	50.970	50.765
5)	MB Aldrin	5.027	3.974	157.1E6	188.9E6	50.603	50.870
6)	B beta-BHC	4.334	3.676	70445128	76200383	49.892	50.291
7)	B delta-BHC	4.574	3.896	150.1E6	197.1E6	54.109	51.231
8)	B Heptachlo...	5.460	4.471	141.3E6	168.9E6	49.481	50.052
9)	A Endosulfan I	5.844	4.834	134.0E6	161.0E6	50.396	50.325
10)	B gamma-Chl...	5.716	4.719	145.7E6	170.5E6	50.938	50.386
11)	B alpha-Chl...	5.793	4.781	143.2E6	166.8E6	49.706	49.818
12)	B 4,4'-DDE	5.975	4.976	127.6E6	156.8E6	51.056	51.144
13)	MA Dieldrin	6.122	5.095	137.2E6	168.5E6	50.012	49.972
14)	MA Endrin	6.352	5.366	118.1E6	146.6E6	48.823	48.477
15)	B Endosulfa...	6.583	5.663	114.3E6	134.0E6	50.635	47.875
16)	A 4,4'-DDD	6.501	5.527	104.1E6	128.7E6	52.480	51.462
17)	MA 4,4'-DDT	6.810	5.773	109.9E6	126.7E6	48.451	48.464
18)	B Endrin al...	6.714	5.842	89050143	101.8E6	49.168	47.839
19)	B Endosulfa...	6.947	6.065	113.8E6	135.2E6	49.055	47.928
20)	A Methoxychlor	7.294	6.351	56151787	63565506	46.545	46.768
21)	B Endrin ke...	7.433	6.561	124.5E6	149.4E6	50.468	48.522
22)	Mirex	7.889	6.733	91803823	106.6E6	48.468	45.061

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\p1120522\
Data File : PL079559.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Dec 2022 10:31
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 06 01:14:28 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112322.M
Quant Title : GC Extractables
QLast Update : Thu Nov 24 00:09:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

