

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL120619\
 Data File : PL054859.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Dec 2019 03:01
 Operator : SG\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 07 04:17:44 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112619.M
 Quant Title : GC Extractables
 QLast Update : Wed Nov 27 01:19:55 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.537	4.051	267.5E6	65608025	22.992	25.710
28)	SA Decachlor...	8.316	9.148	159.3E6	49258369	16.633	20.546
Target Compounds							
7)	B delta-BHC	4.741	0.000	1904793	0	0.117	N.D. #
8)	B Heptachlo...	5.275	5.923	56245398	17676718	3.718	7.076 #
12)	B 4,4'-DDE	5.733	0.000	3334545	0	0.251	N.D. #
14)	MA Endrin	0.000	6.756	0	4004997	N.D.	1.808 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL120619\
Data File : PL054859.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Dec 2019 03:01
Operator : SG\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 07 04:17:44 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112619.M
Quant Title : GC Extractables
QLast Update : Wed Nov 27 01:19:55 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.537 min
Delta R.T.: 0.000 min
Response: 267489216
Conc: 22.99 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.051 min
Delta R.T.: 0.000 min
Response: 65608025
Conc: 25.71 ng/ml




#7 delta-BHC

R.T.: 4.741 min
Delta R.T.: 0.015 min
Response: 1904793
Conc: 0.12 ng/ml



#7 delta-BHC


R.T.: 0.000 min
Exp R.T.: 5.117 min
Response: 0
Conc: N.D.



#8 Heptachlor epoxide


R.T.: 5.275 min
Delta R.T.: 0.005 min
Response: 56245398
Conc: 3.72 ng/ml

Instrument :
ECD_L
ClientSampleId :




#8 Heptachlor epoxide

R.T.: 5.923 min
Delta R.T.: -0.008 min
Response: 17676718
Conc: 7.08 ng/ml




#12 4,4'-DDE

R.T.: 5.733 min
Delta R.T.: 0.013 min
Response: 3334545
Conc: 0.25 ng/ml



#12 4,4'-DDE


R.T.: 0.000 min
Exp R.T.: 6.394 min
Response: 0
Conc: N.D.



#14 Endrin


R.T.: 0.000 min
Exp R.T.: 6.115 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#14 Endrin

R.T.: 6.756 min
Delta R.T.: -0.008 min
Response: 4004997
Conc: 1.81 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.316 min
Delta R.T.: -0.003 min
Response: 159276004
Conc: 16.63 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.148 min
Delta R.T.: -0.003 min
Response: 49258369
Conc: 20.55 ng/ml