

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL120619\
 Data File : PL054819.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Dec 2019 11:35
 Operator : SG\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 07 00:03:48 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112619.M
 Quant Title : GC Extractables
 QLast Update : Wed Nov 27 01:19:55 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.540	4.050	583.1E6	135.1E6	50.120	52.961
28)	SA Decachlor...	8.320	9.150	405.4E6	109.5E6	42.340	45.685
Target Compounds							
2)	A alpha-BHC	3.983	4.442	966.3E6	195.0E6	50.902	55.030
3)	MA gamma-BHC...	4.270	4.727	880.3E6	176.7E6	51.641	54.202
4)	MA Heptachlor	4.576	5.237	866.2E6	154.6E6	49.120	51.218
5)	MB Aldrin	4.826	5.543	834.6E6	149.1E6	51.205	53.511
6)	B beta-BHC	4.521	4.898	370.3E6	77247910	49.922	54.855
7)	B delta-BHC	4.729	5.116	838.5E6	161.2E6	51.582	60.478
8)	B Heptachlo...	5.272	5.930	772.1E6	139.5E6	51.044	55.853
9)	A Endosulfan I	5.614	6.287	700.4E6	127.7E6	51.360	51.621
10)	B gamma-Chl...	5.500	6.164	778.4E6	138.5E6	51.333	52.576
11)	B alpha-Chl...	5.559	6.237	746.7E6	136.0E6	50.639	51.802
12)	B 4,4'-DDE	5.722	6.393	674.3E6	127.9E6	50.679	51.665
13)	MA Dieldrin	5.858	6.545	737.3E6	130.7E6	51.992	51.333
14)	MA Endrin	6.117	6.764	602.9E6	102.4E6	47.344	46.237
15)	B Endosulfa...	6.390	6.974	590.0E6	113.9E6	48.761	48.443
16)	A 4,4'-DDD	6.239	6.885	539.9E6	103.7E6	55.274	53.992
17)	MA 4,4'-DDT	6.481	7.185	446.8E6	86384393	42.649	43.599
18)	B Endrin al...	6.558	7.098	462.6E6	98257288	46.713	50.126
19)	B Endosulfa...	6.772	7.321	487.3E6	103.4E6	45.508	48.784
20)	A Methoxychlor	7.026	7.645	204.1E6	47239488	40.714	41.413
21)	B Endrin ke...	7.265	7.796	528.9E6	114.2E6	48.659	50.716
22)	Mirex	7.457	8.252	369.2E6	80643051	45.471	46.660

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL120619\
Data File : PL054819.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Dec 2019 11:35
Operator : SG\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 07 00:03:48 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112619.M
Quant Title : GC Extractables
QLast Update : Wed Nov 27 01:19:55 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

