

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL120619\
 Data File : PL054824.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Dec 2019 13:09
 Operator : SG\AJ
 Sample : K6162-01MS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TP1MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 07 00:05:04 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112619.M
 Quant Title : GC Extractables
 QLast Update : Wed Nov 27 01:19:55 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.536	4.050	203.0E6	49576846	17.445	19.428
28)	SA Decachlor...	8.317	9.151	134.7E6	37612971	14.065	15.689
Target Compounds							
2)	A alpha-BHC	3.979	4.442	992.2E6	208.4E6	52.264	58.827
3)	MA gamma-BHC...	4.266	4.727	893.7E6	193.4E6	52.427	59.301
4)	MA Heptachlor	4.572	5.237	923.3E6	173.8E6	52.358	57.598
5)	MB Aldrin	4.823	5.542	854.8E6	168.1E6	52.442	60.329
6)	B beta-BHC	4.517	4.898	391.8E6	80286152	52.831	57.012
7)	B delta-BHC	4.724	5.116	822.5E6	147.6E6	50.602	55.377
8)	B Heptachlo...	5.269	5.929	781.3E6	144.4E6	51.650	57.796
9)	A Endosulfan I	5.611	6.287	663.4E6	140.2E6	48.645	56.680
10)	B gamma-Chl...	5.497	6.163	797.4E6	154.5E6	52.582	58.664
11)	B alpha-Chl...	5.555	6.236	747.6E6	155.1E6	50.697	59.073
12)	B 4,4'-DDE	5.718	6.392	719.0E6	142.7E6	54.038	57.659
13)	MA Dieldrin	5.855	6.544	729.5E6	143.8E6	51.442	56.510
14)	MA Endrin	6.114	6.763	628.7E6	119.8E6	49.376	54.093
15)	B Endosulfa...	6.387	6.972	595.9E6	131.1E6	49.244	55.747
16)	A 4,4'-DDD	6.236	6.884	505.2E6	117.5E6	51.725	61.207
17)	MA 4,4'-DDT	6.478	7.184	515.6E6	110.1E6	49.219	55.593
18)	B Endrin al...	6.555	7.097	458.3E6	106.7E6	46.283	54.451
19)	B Endosulfa...	6.768	7.321	497.3E6	114.9E6	46.435	54.201
20)	A Methoxychlor	7.022	7.644	238.0E6	57695418	47.496	50.579
21)	B Endrin ke...	7.262	7.795	502.8E6	127.3E6	46.264	56.533
22)	Mirex	7.454	8.252	389.6E6	96955353	47.988	56.098

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL120619\
Data File : PL054824.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Dec 2019 13:09
Operator : SG\AJ
Sample : K6162-01MS
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
TP1MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 07 00:05:04 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112619.M
Quant Title : GC Extractables
QLast Update : Wed Nov 27 01:19:55 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

