

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL120619\
 Data File : PL054846.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Dec 2019 23:01
 Operator : SG\AJ
 Sample : K6147-15MS
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 P001-WC004MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 07 00:09:45 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112619.M
 Quant Title : GC Extractables
 QLast Update : Wed Nov 27 01:19:55 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.537	4.051	192.3E6	53676363	16.526	21.035 #
28)	SA Decachlor...	8.317	9.149	186.0E6	50120100	19.421	20.906
Target Compounds							
2)	A alpha-BHC	3.980	4.442	935.6E6	222.3E6	49.282	62.755 #
3)	MA gamma-BHC...	4.267	4.728	860.4E6	205.0E6	50.474	62.862
4)	MA Heptachlor	4.572	5.238	842.6E6	177.9E6	47.784	58.937
5)	MB Aldrin	4.823	5.543	726.4E6	167.3E6	44.563	60.046 #
6)	B beta-BHC	4.517	4.898	361.9E6	85720195	48.790	60.871
7)	B delta-BHC	4.725	5.116	797.0E6	186.9E6	49.029	70.091 #
8)	B Heptachlo...	5.269	5.930	781.1E6	164.9E6	51.638	65.996 #
9)	A Endosulfan I	5.610	6.287	692.7E6	150.9E6	50.794	61.012
10)	B gamma-Chl...	5.497	6.164	782.3E6	163.1E6	51.590	61.922
11)	B alpha-Chl...	5.555	6.236	748.6E6	160.2E6	50.769	60.998
12)	B 4,4'-DDE	5.718	6.393	699.4E6	154.0E6	52.566	62.224
13)	MA Dieldrin	5.855	6.545	747.2E6	159.5E6	52.692	62.677
14)	MA Endrin	6.113	6.763	650.7E6	130.5E6	51.104	58.902
15)	B Endosulfa...	6.386	6.973	600.7E6	141.8E6	49.644	60.332
16)	A 4,4'-DDD	6.235	6.885	562.9E6	129.3E6	57.631	67.319
17)	MA 4,4'-DDT	6.478	7.184	491.5E6	110.1E6	46.911	55.576
18)	B Endrin al...	6.554	7.097	419.3E6	100.4E6	42.346	51.223
19)	B Endosulfa...	6.768	7.321	537.4E6	127.0E6	50.179	59.895
20)	A Methoxychlor	7.023	7.643	252.2E6	59205213	50.323	51.903
21)	B Endrin ke...	7.262	7.795	617.2E6	143.0E6	56.781	63.542
22)	Mirex	7.454	8.252	438.4E6	108.9E6	54.000	63.028

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL120619\
Data File : PL054846.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Dec 2019 23:01
Operator : SG\AJ
Sample : K6147-15MS
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
P001-WC004MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 07 00:09:45 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112619.M
Quant Title : GC Extractables
QLast Update : Wed Nov 27 01:19:55 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

