

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL120718\
 Data File : PL042872.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Dec 2018 11:36
 Operator : AJ\SJ
 Sample : J6188-02
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 07 13:08:49 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL120618.M
 Quant Title : GC Extractables
 QLast Update : Fri Dec 07 00:20:08 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.465	4.099	118.7E6	61240281	19.357	20.662
28)	SA Decachlor...	8.199	9.226	117.7E6	49114947	18.017	18.829
Target Compounds							
6)	B beta-BHC	4.397f	0.000	8381624	0	2.272	N.D. #
7)	B delta-BHC	0.000	5.145f	0	8691382	N.D.	2.022 #
8)	B Heptachlo...	0.000	5.992	0	1827432	N.D.	0.482 #
14)	MA Endrin	5.995	0.000	3545583	0	0.492	N.D. #
19)	B Endosulfa...	0.000	7.451f	0	2755135	N.D.	0.916 #
23)	Chlordane-1	4.361	5.145	2508413	8691382	9.335	61.496 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL120718\
Data File : PL042872.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Dec 2018 11:36
Operator : AJ\SJ
Sample : J6188-02
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 07 13:08:49 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL120618.M
Quant Title : GC Extractables
QLast Update : Fri Dec 07 00:20:08 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.465 min
Delta R.T.: 0.000 min
Response: 118716176
Conc: 19.36 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.099 min
Delta R.T.: 0.000 min
Response: 61240281
Conc: 20.66 ng/ml




#6 beta-BHC

R.T.: 4.397 min
Delta R.T.: -0.043 min
Response: 8381624
Conc: 2.27 ng/ml



#6 beta-BHC


R.T.: 0.000 min
Exp R.T. : 4.952 min
Response: 0
Conc: N.D.



#7 delta-BHC


R.T.: 0.000 min
Exp R.T. : 4.643 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#7 delta-BHC

R.T.: 5.145 min
Delta R.T.: -0.027 min
Response: 8691382
Conc: 2.02 ng/ml




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T. : 5.171 min
Response: 0
Conc: N.D.



#8 Heptachlor epoxide


R.T.: 5.992 min
Delta R.T.: -0.002 min
Response: 1827432
Conc: 0.48 ng/ml



#14 Endrin


R.T.: 5.995 min
Delta R.T.: -0.012 min
Response: 3545583
Conc: 0.49 ng/ml

Instrument :
ECD_L
ClientSampleId :




#14 Endrin

R.T.: 0.000 min
Exp R.T. : 6.832 min
Response: 0
Conc: N.D.




#19 Endosulfan Sulfate

R.T.: 0.000 min
Exp R.T. : 6.663 min
Response: 0
Conc: N.D.



#19 Endosulfan Sulfate


R.T.: 7.451 min
Delta R.T.: 0.063 min
Response: 2755135
Conc: 0.92 ng/ml



#23 Chlordane-1


R.T.: 4.361 min
Delta R.T.: 0.029 min
Response: 2508413
Conc: 9.34 ng/ml

Instrument :
ECD_L
ClientSampleId :




#23 Chlordane-1

R.T.: 5.145 min
Delta R.T.: 0.037 min
Response: 8691382
Conc: 61.50 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.199 min
Delta R.T.: -0.002 min
Response: 117652522
Conc: 18.02 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.226 min
Delta R.T.: 0.000 min
Response: 49114947
Conc: 18.83 ng/ml