

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL120920\
 Data File : PL063410.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Dec 2020 09:45
 Operator : DD\AJ
 Sample : PB133459BL
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 09 14:02:34 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112320.M
 Quant Title : GC Extractables
 QLast Update : Tue Nov 24 01:53:58 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.417	3.952	38481053	74554692	22.595	22.843
28)	SA Decachlor...	8.137	8.999	42084422	57507885	22.536	24.171
Target Compounds							
8)	B Heptachlo...	0.000	5.814	0	33661033	N.D.	9.663 #
14)	MA Endrin	5.947	0.000	16113104	0	9.595	N.D. #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL120920\
Data File : PL063410.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Dec 2020 09:45
Operator : DD\AJ
Sample : PB133459BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 09 14:02:34 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112320.M
Quant Title : GC Extractables
QLast Update : Tue Nov 24 01:53:58 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.417 min
Delta R.T.: 0.007 min
Response: 38481053
Conc: 22.60 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.952 min
Delta R.T.: 0.002 min
Response: 74554692
Conc: 22.84 ng/ml




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T. : 5.101 min
Response: 0
Conc: N.D.



#8 Heptachlor epoxide


R.T.: 5.814 min
Delta R.T.: 0.003 min
Response: 33661033
Conc: 9.66 ng/ml



#14 Endrin


R.T.: 5.947 min
Delta R.T.: 0.015 min
Response: 16113104
Conc: 9.59 ng/ml

Instrument :
ECD_L
ClientSampleId :




#14 Endrin

R.T.: 0.000 min
Exp R.T. : 6.642 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 8.137 min
Delta R.T.: 0.011 min
Response: 42084422
Conc: 22.54 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.999 min
Delta R.T.: 0.009 min
Response: 57507885
Conc: 24.17 ng/ml