

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL120921\
 Data File : PL071354.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Dec 2021 15:07
 Operator : AR\AJ
 Sample : PB141267BL
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 10 00:41:32 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL113021.M
 Quant Title : GC Extractables
 QLast Update : Tue Nov 30 16:26:31 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.569	5.501	9563357	23296111	20.361	18.564
28)	SA Decachlor...	10.330	11.395	13564891	22624676	18.734	16.277
Target Compounds							
7)	B delta-BHC	0.000	7.026f	0	879387	N.D.	0.517 #
20)	A Methoxychlor	9.015	0.000	156144	0	0.432	N.D. #
21)	B Endrin ke...	9.237	0.000	25230	0	0.035	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL120921\
Data File : PL071354.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Dec 2021 15:07
Operator : AR\AJ
Sample : PB141267BL
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 10 00:41:32 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL113021.M
Quant Title : GC Extractables
QLast Update : Tue Nov 30 16:26:31 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 4.569 min
Delta R.T.: 0.003 min
Response: 9563357
Conc: 20.36 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 5.501 min
Delta R.T.: 0.000 min
Response: 23296111
Conc: 18.56 ng/ml




#7 delta-BHC

R.T.: 0.000 min
Exp R.T. : 6.337 min
Response: 0
Conc: N.D.



#7 delta-BHC


R.T.: 7.026 min
Delta R.T.: 0.025 min
Response: 879387
Conc: 0.52 ng/ml



#20 Methoxychlor


R.T.: 9.015 min
Delta R.T.: 0.008 min
Response: 156144
Conc: 0.43 ng/ml

Instrument :
ECD_L
ClientSampleId :




#20 Methoxychlor

R.T.: 0.000 min
Exp R.T. : 9.843 min
Response: 0
Conc: N.D.




#21 Endrin ketone

R.T.: 9.237 min
Delta R.T.: 0.005 min
Response: 25230
Conc: 0.04 ng/ml



#21 Endrin ketone


R.T.: 0.000 min
Exp R.T. : 9.997 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 10.330 min
Delta R.T.: 0.002 min
Response: 13564891
Conc: 18.73 ng/ml

Instrument :
ECD_L
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 11.395 min
Delta R.T.: 0.003 min
Response: 22624676
Conc: 16.28 ng/ml