

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL120925\
 Data File : PL097800.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Dec 2025 10:30
 Operator : AR\AJ
 Sample : PSTDICC075
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC075

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 09 11:06:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL120925.M
 Quant Title : GC Extractables
 QLast Update : Tue Dec 09 11:02:02 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.498	2.800	378.2E6	523.9E6	74.020	73.664
28)	SA Decachlor...	8.957	7.947	278.3E6	471.0E6	75.830	73.445
Target Compounds							
2)	A alpha-BHC	3.945	3.302	674.5E6	894.4E6	75.235	73.751
3)	MA gamma-BHC...	4.273	3.633	622.0E6	827.4E6	74.419	73.668
4)	MA Heptachlor	4.862	3.979	613.3E6	785.3E6	75.448	73.642
5)	MB Aldrin	5.201	4.260	563.8E6	750.2E6	75.060	73.629
6)	B beta-BHC	4.462	3.929	248.7E6	342.0E6	74.059	73.587
7)	B delta-BHC	4.707	4.160	607.7E6	826.9E6	74.545	73.718
8)	B Heptachlo...	5.621	4.760	505.3E6	674.3E6	74.858	73.639
9)	A Endosulfan I	6.002	5.130	447.4E6	668.0E6	74.886	73.157
10)	B gamma-Chl...	5.874	5.012	501.9E6	744.0E6	75.526	73.452
11)	B alpha-Chl...	5.954	5.076	491.0E6	730.2E6	74.099	73.524
12)	B 4,4'-DDE	6.126	5.264	456.8E6	682.3E6	75.459	73.720
13)	MA Dieldrin	6.274	5.394	479.9E6	714.5E6	75.773	73.713
14)	MA Endrin	6.500	5.668	411.8E6	637.9E6	76.141	73.802
15)	B Endosulfa...	6.714	5.960	389.2E6	616.3E6	75.937	73.724
16)	A 4,4'-DDD	6.635	5.816	360.7E6	570.4E6	75.664	73.303
17)	MA 4,4'-DDT	6.948	6.068	379.7E6	597.5E6	75.593	73.812
18)	B Endrin al...	6.843	6.139	296.0E6	463.3E6	75.363	73.662
19)	B Endosulfa...	7.076	6.362	371.4E6	605.3E6	75.289	73.715
20)	A Methoxychlor	7.421	6.640	191.9E6	307.9E6	75.493	73.795
21)	B Endrin ke...	7.554	6.865	386.9E6	662.4E6	76.071	73.689
22)	Mirex	8.030	7.053	256.1E6	462.9E6	75.989	73.506

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL120925\
Data File : PL097800.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Dec 2025 10:30
Operator : AR\AJ
Sample : PSTDICC075
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC075

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 09 11:06:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL120925.M
Quant Title : GC Extractables
QLast Update : Tue Dec 09 11:02:02 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

