

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL1121019\
 Data File : PL054891.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Dec 2019 10:19
 Operator : SG\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 12/11/2019 10:03:57 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 11 00:37:42 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112619.M
 Quant Title : GC Extractables
 QLast Update : Wed Nov 27 01:19:55 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.542	4.051	586.7E6	137.7E6	50.428	53.954
28)	SA Decachlor...	8.325	9.154	460.8E6	115.2E6	48.117	48.059
Target Compounds							
2)	A alpha-BHC	3.986	4.442	960.2E6	197.1E6	50.578	55.624
3)	MA gamma-BHC...	4.273	4.728	884.5E6	180.2E6	51.888	55.256
4)	MA Heptachlor	4.578	5.238	908.7E6	164.9E6	51.534	54.643
5)	MB Aldrin	4.830	5.544	852.7E6	152.6E6	52.313	54.778
6)	B beta-BHC	4.523	4.898	369.5E6	82271047	49.820	58.422m
7)	B delta-BHC	4.731	5.117	854.4E6	157.8E6	52.564	59.194
8)	B Heptachlor...	5.275	5.931	829.2E6	136.4E6	54.817	54.595
9)	A Endosulfan I	5.617	6.289	744.8E6	130.2E6	54.620	52.635
10)	B gamma-Chl...	5.504	6.166	828.6E6	141.6E6	54.641	53.764
11)	B alpha-Chl...	5.562	6.238	784.7E6	139.2E6	53.212	53.007
12)	B 4,4'-DDE	5.725	6.395	694.9E6	131.3E6	52.230	53.023
13)	MA Dieldrin	5.861	6.547	771.7E6	135.0E6	54.421	53.042
14)	MA Endrin	6.120	6.766	682.8E6	115.3E6	53.619	52.029
15)	B Endosulfa...	6.393	6.975	614.7E6	120.3E6	50.802	51.165
16)	A 4,4'-DDD	6.243	6.887	527.6E6	103.3E6	54.022	53.812
17)	MA 4,4'-DDT	6.485	7.187	569.0E6	105.6E6	54.318	53.299
18)	B Endrin al...	6.561	7.100	505.6E6	100.9E6	51.056	51.464
19)	B Endosulfa...	6.774	7.324	547.1E6	109.1E6	51.092	51.448
20)	A Methoxychlor	7.029	7.647	275.4E6	57831719	54.948	50.699
21)	B Endrin ke...	7.269	7.798	575.1E6	114.3E6	52.909	50.767
22)	Mirex	7.460	8.256	393.6E6	85744435	48.479	49.611

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL121019\
 Data File : PL054891.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Dec 2019 10:19
 Operator : SG\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 12/11/2019 10:03:57 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 11 00:37:42 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112619.M
 Quant Title : GC Extractables
 QLast Update : Wed Nov 27 01:19:55 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

