

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL121024\
 Data File : PL093352.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Dec 2024 13:53
 Operator : AR\AJ
 Sample : P5216-01MS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 BP-G-6MS

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 12/11/2024
 Supervised By :Ankita Jodhani 12/11/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 11 01:10:30 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112524.M
 Quant Title : GC Extractables
 QLast Update : Mon Nov 25 15:18:43 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.537	2.776	44848798	50056047	17.256m	17.356
2) SA Decachlor...	9.054	7.913	31345171	44823855	18.030	15.692
Target Compounds						
2) A alpha-BHC	3.994	3.278	175.1E6	219.0E6	49.058	51.299
3) MA gamma-BHC...	4.327	3.608	164.8E6	207.6E6	48.762	50.154
4) MA Heptachlor	4.915	3.947	153.0E6	216.3E6	50.055	53.469
5) MB Aldrin	5.256	4.226	144.8E6	203.5E6	48.152m	51.133
6) B beta-BHC	4.525	3.908	74329708	91621029	49.234	51.505
7) B delta-BHC	4.772	4.137	149.8E6	204.7E6	45.275	47.984
8) B Heptachlo...	5.684	4.729	134.3E6	189.2E6	48.380	51.959
9) A Endosulfan I	6.069	5.099	122.8E6	172.6E6	50.445	51.645
10) B gamma-Chl...	5.939	4.978	127.7E6	196.9E6	49.554	53.135m
11) B alpha-Chl...	6.019	5.043	131.7E6	194.9E6	50.847	53.689
12) B 4,4'-DDE	6.192	5.231	114.1E6	190.0E6	48.769	53.071
13) MA Dieldrin	6.344	5.363	128.4E6	195.9E6	50.081	53.159
14) MA Endrin	6.574	5.637	112.2E6	175.4E6	53.470	54.998m
15) B Endosulfa...	6.792	5.933	102.3E6	166.1E6	46.895m	52.407
16) A 4,4'-DDD	6.708	5.787	93357766	140.4E6	50.951m	50.089
17) MA 4,4'-DDT	7.023	6.037	102.6E6	157.5E6	53.221	53.173
18) B Endrin al...	6.924	6.113	88219838	128.9E6	48.832	49.148
19) B Endosulfa...	7.158	6.335	105.3E6	153.9E6	50.798	50.639
20) A Methoxychlor	7.499	6.612	54320389	79486299	51.987	52.056
21) B Endrin ke...	7.643	6.841	118.4E6	169.0E6	52.166	50.348
22) Mirex	8.116	7.021	89596676	129.0E6	49.604	48.014

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL121024\
Data File : PL093352.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Dec 2024 13:53
Operator : AR\AJ
Sample : P5216-01MS
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

BP-G-6MS

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 12/11/2024

Supervised By :Ankita Jodhani 12/11/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 11 01:10:30 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112524.M
Quant Title : GC Extractables
QLast Update : Mon Nov 25 15:18:43 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

