

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL1121219\
 Data File : PL054955.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Dec 2019 09:37
 Operator : SG\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 13 03:29:37 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112619.M
 Quant Title : GC Extractables
 QLast Update : Wed Nov 27 01:19:55 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.535	4.047	273.3E6	64952988	23.488	25.454
28)	SA Decachlor...	8.311	9.139	212.0E6	53563022	22.137	22.342
Target Compounds							
8)	B Heptachlo...	0.000	5.917	0	6663810	N.D.	2.667 #
12)	B 4,4'-DDE	5.735f	0.000	17930818	0	1.348	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL1121219\
Data File : PL054955.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Dec 2019 09:37
Operator : SG\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 13 03:29:37 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112619.M
Quant Title : GC Extractables
QLast Update : Wed Nov 27 01:19:55 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.535 min
Delta R.T.: -0.002 min
Response: 273266610
Conc: 23.49 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 4.047 min
Delta R.T.: -0.004 min
Response: 64952988
Conc: 25.45 ng/ml




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T. : 5.270 min
Response: 0
Conc: N.D.



#8 Heptachlor epoxide


R.T.: 5.917 min
Delta R.T.: -0.013 min
Response: 6663810
Conc: 2.67 ng/ml



#12 4,4'-DDE


R.T.: 5.735 min
Delta R.T.: 0.015 min
Response: 17930818
Conc: 1.35 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#12 4,4'-DDE

R.T.: 0.000 min
Exp R.T. : 6.394 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 8.311 min
Delta R.T.: -0.009 min
Response: 211988263
Conc: 22.14 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.139 min
Delta R.T.: -0.013 min
Response: 53563022
Conc: 22.34 ng/ml