

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL121220\
 Data File : PL063504.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Dec 2020 13:20
 Operator : DD\AJ
 Sample : L4984-02
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 11 16:40:01 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL121020.M
 Quant Title : GC Extractables
 QLast Update : Thu Dec 10 15:57:40 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.410	3.949	35191935	51396228	20.680	16.399
28) SA Decachlor...	8.125	8.989	35444139	51448825	17.600	20.430
Target Compounds						
7) B delta-BHC	0.000	4.983f	0	12718602	N.D.	3.142 #
8) B Heptachlo...	0.000	5.824	0	9860295	N.D.	2.722 #
14) MA Endrin	5.935	0.000	4429839	0	2.639	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL121220\
Data File : PL063504.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Dec 2020 13:20
Operator : DD\AJ
Sample : L4984-02
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 11 16:40:01 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL121020.M
Quant Title : GC Extractables
QLast Update : Thu Dec 10 15:57:40 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.410 min
Delta R.T.: 0.000 min
Response: 35191935
Conc: 20.68 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.949 min
Delta R.T.: -0.001 min
Response: 51396228
Conc: 16.40 ng/ml




#7 delta-BHC

R.T.: 0.000 min
Exp R.T. : 4.572 min
Response: 0
Conc: N.D.



#7 delta-BHC


R.T.: 4.983 min
Delta R.T.: -0.023 min
Response: 12718602
Conc: 3.14 ng/ml



#8 Heptachlor epoxide


R.T.: 0.000 min
Exp R.T.: 5.102 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#8 Heptachlor epoxide

R.T.: 5.824 min
Delta R.T.: 0.012 min
Response: 9860295
Conc: 2.72 ng/ml




#14 Endrin

R.T.: 5.935 min
Delta R.T.: 0.002 min
Response: 4429839
Conc: 2.64 ng/ml



#14 Endrin

R.T.: 0.000 min
Exp R.T.: 6.643 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl


R.T.: 8.125 min

Delta R.T.: -0.003 min

Response: 35444139

Conc: 17.60 ng/ml

Instrument :
ECD_L
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 8.989 min

Delta R.T.: -0.004 min

Response: 51448825

Conc: 20.43 ng/ml