

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL121220\
 Data File : PL063517.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Dec 2020 17:22
 Operator : DD\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 12 00:42:55 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL121020.M
 Quant Title : GC Extractables
 QLast Update : Thu Dec 10 15:57:40 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.410	3.949	94151490	160.2E6	55.328	51.126
28)	SA Decachlor...	8.125	8.989	97831452	139.9E6	48.579	55.545
Target Compounds							
2)	A alpha-BHC	3.842	4.336	138.8E6	249.7E6	55.391	52.754
3)	MA gamma-BHC...	4.121	4.619	125.8E6	229.7E6	55.199	51.529
4)	MA Heptachlor	4.418	5.124	111.2E6	200.9E6	51.799	48.933
5)	MB Aldrin	4.662	5.427	106.7E6	199.3E6	55.982	51.517
6)	B beta-BHC	4.371	4.791	52703219	94452817	54.278	47.548
7)	B delta-BHC	4.572	5.005	118.2E6	207.5E6	54.302	51.261
8)	B Heptachlo...	5.101	5.812	102.3E6	183.3E6	54.981	50.615
9)	A Endosulfan I	5.436	6.167	96181421	170.6E6	55.107	51.123
10)	B gamma-Chl...	5.327	6.046	102.6E6	181.2E6	55.160	51.053
11)	B alpha-Chl...	5.384	6.118	101.8E6	178.0E6	54.769	50.226
12)	B 4,4'-DDE	5.548	6.277	95259119	176.1E6	56.086	50.282
13)	MA Dieldrin	5.677	6.424	101.7E6	176.2E6	55.872	51.280
14)	MA Endrin	5.932	6.641	76537321	113.3E6	45.602	38.751
15)	B Endosulfa...	6.204	6.851	92795145	157.9E6	52.433	49.576
16)	A 4,4'-DDD	6.061	6.767	88360955	149.1E6	58.865	55.012
17)	MA 4,4'-DDT	6.299	7.065	62641894	107.9E6	45.649	42.200
18)	B Endrin al...	6.371	6.975	83895998	138.5E6	54.732	54.286
19)	B Endosulfa...	6.583	7.198	90296598	137.4E6	52.653	49.381
20)	A Methoxychlor	6.843	7.525	39972475	59151897	44.922	42.413
21)	B Endrin ke...	7.069	7.668	115.4E6	156.1E6	57.786	53.525
22)	Mirex	7.255	8.120	80602405	100.3E6	50.767	50.360

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL121220\
Data File : PL063517.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Dec 2020 17:22
Operator : DD\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 12 00:42:55 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL121020.M
Quant Title : GC Extractables
QLast Update : Thu Dec 10 15:57:40 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

