

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL121323\
 Data File : PL087147.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Dec 2023 19:36
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 13 23:36:49 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL121223.M
 Quant Title : GC Extractables
 QLast Update : Tue Dec 12 17:22:26 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.456	2.704	176.8E6	49513731	48.451	50.133
28)	SA Decachlor...	9.050	7.900	159.1E6	53272538	52.563	52.005
Target Compounds							
2)	A alpha-BHC	3.914	3.211	258.6E6	68113565	49.221	51.579
3)	MA gamma-BHC...	4.250	3.542	231.4E6	67063695	47.108	51.220
4)	MA Heptachlor	4.845	3.886	208.2E6	63710768	41.065	49.070
5)	MB Aldrin	5.188	4.167	195.2E6	61531153	40.704	50.150
6)	B beta-BHC	4.450	3.844	94687182	30194628	45.069	50.612
7)	B delta-BHC	4.697	4.074	202.0E6	67172545	44.156	52.346
8)	B Heptachlo...	5.621	4.675	183.5E6	57737292	43.285	50.035
9)	A Endosulfan I	6.009	5.046	183.9E6	54727764	45.581	45.019
10)	B gamma-Chl...	5.881	4.929	195.8E6	59281001	45.632	48.440
11)	B alpha-Chl...	5.961	4.993	195.3E6	59151357	45.656	47.016
12)	B 4,4'-DDE	6.146	5.192	158.1E6	51623948	48.313	49.664
13)	MA Dieldrin	6.288	5.314	189.5E6	56226930	46.505	48.320
14)	MA Endrin	6.520	5.589	150.8E6	49620476	45.264	48.532
15)	B Endosulfa...	6.742	5.887	162.8E6	48761795	44.934	49.818
16)	A 4,4'-DDD	6.669	5.750	130.9E6	43850421	53.797	52.883
17)	MA 4,4'-DDT	6.988	6.003	136.9E6	44198150	48.460	50.380
18)	B Endrin al...	6.875	6.070	117.6E6	37983364	49.974	47.670
19)	B Endosulfa...	7.113	6.295	162.2E6	48852072	49.340	52.287
20)	A Methoxychlor	7.480	6.590	72335421	26178657	49.463	48.014
21)	B Endrin ke...	7.601	6.800	166.8E6	57172208	51.415	53.196
22)	Mirex	8.079	6.983	143.8E6	47166661	48.860	50.190

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL121232\
Data File : PL087147.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Dec 2023 19:36
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 13 23:36:49 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL121223.M
Quant Title : GC Extractables
QLast Update : Tue Dec 12 17:22:26 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

