

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL121418\
 Data File : PL043160.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Dec 2018 19:08
 Operator : AJ\SJ
 Sample : PB115700BL
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 14 22:31:09 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL120618.M
 Quant Title : GC Extractables
 QLast Update : Fri Dec 07 00:20:08 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.464	4.098	124.5E6	65495612	20.304	22.098
28)	SA Decachlor...	8.197	9.223	135.7E6	55244864	20.782	21.179
Target Compounds							
2)	A alpha-BHC	0.000	4.451f	0	417071	N.D.	0.092 #
6)	B beta-BHC	4.396f	0.000	3256660	0	0.883	N.D. #
7)	B delta-BHC	0.000	5.156	0	5334429	N.D.	1.241 #
8)	B Heptachlo...	0.000	5.992	0	2905571	N.D.	0.766 #
23)	Chlordane-1	4.360	5.156	1241046	5334429	4.619	37.744 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL121418\
Data File : PL043160.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Dec 2018 19:08
Operator : AJ\SJ
Sample : PB115700BL
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 14 22:31:09 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL120618.M
Quant Title : GC Extractables
QLast Update : Fri Dec 07 00:20:08 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.464 min
Delta R.T.: 0.000 min
Response: 124524030
Conc: 20.30 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.098 min
Delta R.T.: -0.001 min
Response: 65495612
Conc: 22.10 ng/ml




#2 alpha-BHC

R.T.: 0.000 min
Exp R.T. : 3.904 min
Response: 0
Conc: N.D.



#2 alpha-BHC


R.T.: 4.451 min
Delta R.T.: -0.044 min
Response: 417071
Conc: 0.09 ng/ml



#6 beta-BHC


R.T.: 4.396 min
Delta R.T.: -0.045 min
Response: 3256660
Conc: 0.88 ng/ml

Instrument :
ECD_L
ClientSampleId :




#6 beta-BHC

R.T.: 0.000 min
Exp R.T. : 4.952 min
Response: 0
Conc: N.D.




#7 delta-BHC

R.T.: 0.000 min
Exp R.T. : 4.643 min
Response: 0
Conc: N.D.



#7 delta-BHC


R.T.: 5.156 min
Delta R.T.: -0.016 min
Response: 5334429
Conc: 1.24 ng/ml



#8 Heptachlor epoxide


R.T.: 0.000 min
Exp R.T.: 5.171 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#8 Heptachlor epoxide

R.T.: 5.992 min
Delta R.T.: -0.003 min
Response: 2905571
Conc: 0.77 ng/ml




#23 Chlordane-1

R.T.: 4.360 min
Delta R.T.: 0.028 min
Response: 1241046
Conc: 4.62 ng/ml



#23 Chlordane-1

R.T.: 5.156 min
Delta R.T.: 0.048 min
Response: 5334429
Conc: 37.74 ng/ml



#28 Decachlorobiphenyl


R.T.: 8.197 min

Delta R.T.: -0.004 min

Response: 135705797

Conc: 20.78 ng/ml

Instrument :
ECD_L
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 9.223 min

Delta R.T.: -0.004 min

Response: 55244864

Conc: 21.18 ng/ml