

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL121720\
 Data File : PL063675.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Dec 2020 18:02
 Operator : DD\AJ
 Sample : L5113-02
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 SS1

Manual Integrations
 APPROVED

Ankita
 12/18/2020 3:20:35 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 18 01:09:23 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL121620.M
 Quant Title : GC Extractables
 QLast Update : Thu Dec 17 09:38:50 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.479	4.147	17025636	26934235	8.370	8.444
28)	SA Decachlor...	8.224	9.335	17838269	20269820	8.165	7.566
Target Compounds							
11)	B alpha-Chl...	5.473	6.373	3831187	12945429	1.388m	3.717m#
12)	B 4,4'-DDE	5.640	6.529	11890827	18061181	4.836	5.594m
17)	MA 4,4'-DDT	6.391	7.329	16060270	11101084	8.275	4.567 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL121720\
Data File : PL063675.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Dec 2020 18:02
Operator : DD\AJ
Sample : L5113-02
Misc :
ALS Vial : 26 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
SS1


Manual Integrations
APPROVED

Ankita
12/18/2020 3:20:35 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 18 01:09:23 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL121620.M
Quant Title : GC Extractables
QLast Update : Thu Dec 17 09:38:50 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.479 min
Delta R.T.: 0.000 min
Response: 17025636
Conc: 8.37 ng/ml

Instrument :
ECD_L
ClientSampleId :
SS1


Manual Integrations
APPROVED

Ankita
12/18/2020 3:20:35 PM




#1 Tetrachloro-m-xylene

R.T.: 4.147 min
Delta R.T.: 0.000 min
Response: 26934235
Conc: 8.44 ng/ml




#11 alpha-Chlordane

R.T.: 5.473 min
Delta R.T.: -0.004 min
Response: 3831187
Conc: 1.39 ng/ml m



#11 alpha-Chlordane

R.T.: 6.373 min
Delta R.T.: -0.003 min
Response: 12945429
Conc: 3.72 ng/ml m




#12 4,4'-DDE

R.T.: 5.640 min
Delta R.T.: -0.003 min
Response: 11890827
Conc: 4.84 ng/ml

Instrument :
ECD_L
ClientSampleId :
SS1


Manual Integrations
APPROVED

Ankita
12/18/2020 3:20:35 PM




#12 4,4'-DDE

R.T.: 6.529 min
Delta R.T.: -0.002 min
Response: 18061181
Conc: 5.59 ng/ml m




#17 4,4'-DDT

R.T.: 6.391 min
Delta R.T.: -0.006 min
Response: 16060270
Conc: 8.27 ng/ml



#17 4,4'-DDT

R.T.: 7.329 min
Delta R.T.: -0.002 min
Response: 11101084
Conc: 4.57 ng/ml




#28 Decachlorobiphenyl

R.T.: 8.224 min
Delta R.T.: -0.005 min
Response: 17838269
Conc: 8.16 ng/ml

Instrument :
ECD_L
ClientSampleId :
SS1

Manual Integrations
APPROVED

Ankita
12/18/2020 3:20:35 PM



#28 Decachlorobiphenyl

R.T.: 9.335 min
Delta R.T.: -0.002 min
Response: 20269820
Conc: 7.57 ng/ml