

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL121819\
 Data File : PL055102.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Dec 2019 23:01
 Operator : SG\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 12/19/2019 11:42:15 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 19 02:17:05 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112619.M
 Quant Title : GC Extractables
 QLast Update : Wed Dec 18 14:58:42 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.536	4.050	523.1E6	134.7E6	44.966	52.804
28)	SA Decachlor...	8.314	9.150	435.7E6	110.6E6	45.494	46.123
Target Compounds							
2)	A alpha-BHC	3.979	4.441	861.4E6	197.7E6	45.373	55.799
3)	MA gamma-BHC...	4.265	4.727	790.1E6	180.2E6	46.351	55.275
4)	MA Heptachlor	4.571	5.236	788.6E6	158.6E6	44.723	52.558
5)	MB Aldrin	4.822	5.541	756.1E6	151.8E6	46.389	54.467
6)	B beta-BHC	4.516	4.897	334.8E6	76677720	45.139	54.450
7)	B delta-BHC	4.724	5.115	758.4E6	164.7E6	46.658	61.757 #
8)	B Heptachlor...	5.266	5.928	701.8E6	142.4E6	46.396m	56.983
9)	A Endosulfan I	5.608	6.286	643.0E6	129.0E6	47.151m	52.154
10)	B gamma-Chl...	5.494	6.163	706.2E6	139.2E6	46.573m	52.840
11)	B alpha-Chl...	5.553	6.235	668.0E6	137.2E6	45.299m	52.264
12)	B 4,4'-DDE	5.717	6.392	628.3E6	128.2E6	47.221	51.789
13)	MA Dieldrin	5.853	6.544	681.4E6	133.3E6	48.050	52.362
14)	MA Endrin	6.112	6.762	577.1E6	100.7E6	45.318	45.475
15)	B Endosulfa...	6.384	6.973	559.7E6	117.9E6	46.256	50.168
16)	A 4,4'-DDD	6.234	6.884	495.8E6	104.8E6	50.762	54.563
17)	MA 4,4'-DDT	6.476	7.184	456.8E6	93605851	43.603	47.244
18)	B Endrin al...	6.553	7.097	460.0E6	99111261	46.448	50.562
19)	B Endosulfa...	6.767	7.321	496.0E6	107.6E6	46.316	50.745
20)	A Methoxychlor	7.020	7.644	219.2E6	49392136	43.738	43.300
21)	B Endrin ke...	7.260	7.795	530.5E6	118.0E6	48.810	52.413
22)	Mirex	7.452	8.252	366.5E6	82668134	45.146	47.831

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL121819\
Data File : PL055102.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Dec 2019 23:01
Operator : SG\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
12/19/2019 11:42:15 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 19 02:17:05 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112619.M
Quant Title : GC Extractables
QLast Update : Wed Dec 18 14:58:42 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

