

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL121820\
 Data File : PL063688.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Dec 2020 10:23
 Operator : DD\AJ
 Sample : PB133369BL
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 18 13:31:26 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL121620.M
 Quant Title : GC Extractables
 QLast Update : Thu Dec 17 09:38:50 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.478	4.147	43546744	69682015	21.409	21.845
28) SA Decachlor...	8.221	9.332	50157186	60458287	22.958	22.566
Target Compounds						
6) B beta-BHC	0.000	5.026	0	6745493	N.D.	1.419 #
14) MA Endrin	6.027	0.000	7374301	0	3.123	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL121820\
Data File : PL063688.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Dec 2020 10:23
Operator : DD\AJ
Sample : PB133369BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 18 13:31:26 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL121620.M
Quant Title : GC Extractables
QLast Update : Thu Dec 17 09:38:50 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.478 min
Delta R.T.: -0.001 min
Response: 43546744
Conc: 21.41 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.147 min
Delta R.T.: -0.001 min
Response: 69682015
Conc: 21.84 ng/ml




#6 beta-BHC

R.T.: 0.000 min
Exp R.T. : 4.449 min
Response: 0
Conc: N.D.



#6 beta-BHC


R.T.: 5.026 min
Delta R.T.: 0.005 min
Response: 6745493
Conc: 1.42 ng/ml



#14 Endrin


R.T.: 6.027 min
Delta R.T.: -0.001 min
Response: 7374301
Conc: 3.12 ng/ml

Instrument :
ECD_L
ClientSampleId :




#14 Endrin

R.T.: 0.000 min
Exp R.T. : 6.915 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 8.221 min
Delta R.T.: -0.008 min
Response: 50157186
Conc: 22.96 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.332 min
Delta R.T.: -0.005 min
Response: 60458287
Conc: 22.57 ng/ml