

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL121919\
 Data File : PL055106.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Dec 2019 09:47
 Operator : SG\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 20 03:52:05 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112619.M
 Quant Title : GC Extractables
 QLast Update : Wed Dec 18 14:58:42 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.535	4.048	517.9E6	127.7E6	44.511	50.027
2)	SA Decachlor...	8.313	9.146	430.8E6	106.6E6	44.988	44.461
Target Compounds							
2)	A alpha-BHC	3.978	4.439	851.4E6	185.0E6	44.849	52.232
3)	MA gamma-BHC...	4.265	4.725	773.1E6	168.8E6	45.354	51.759
4)	MA Heptachlor	4.570	5.234	762.2E6	146.8E6	43.227	48.627
5)	MB Aldrin	4.821	5.539	739.4E6	140.9E6	45.363	50.576
6)	B beta-BHC	4.515	4.896	327.9E6	71906030	44.212	51.061
7)	B delta-BHC	4.723	5.113	741.5E6	154.0E6	45.618	57.761 #
8)	B Heptachlo...	5.266	5.926	693.2E6	128.6E6	45.827	51.473
9)	A Endosulfan I	5.608	6.284	632.4E6	120.4E6	46.371	48.691
10)	B gamma-Chl...	5.494	6.161	696.7E6	130.1E6	45.945	49.388
11)	B alpha-Chl...	5.553	6.233	674.9E6	127.9E6	45.771	48.726
12)	B 4,4'-DDE	5.715	6.390	631.1E6	119.9E6	47.434	48.438
13)	MA Dieldrin	5.852	6.542	664.3E6	124.0E6	46.848	48.727
14)	MA Endrin	6.110	6.760	540.6E6	92627005	42.454	41.811
15)	B Endosulfa...	6.383	6.970	558.8E6	110.4E6	46.185	46.954
16)	A 4,4'-DDD	6.232	6.882	484.0E6	97517784	49.550	50.789
17)	MA 4,4'-DDT	6.474	7.181	446.4E6	85845249	42.608	43.327
18)	B Endrin al...	6.551	7.094	455.5E6	94083392	45.993	47.997
19)	B Endosulfa...	6.764	7.318	484.5E6	100.5E6	45.242	47.376
20)	A Methoxychlor	7.019	7.641	209.8E6	45775236	41.855	40.129
21)	B Endrin ke...	7.258	7.792	535.3E6	112.0E6	49.252	49.761
22)	Mirex	7.450	8.249	356.5E6	79518054	43.915	46.009

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL121919\
Data File : PL055106.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Dec 2019 09:47
Operator : SG\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 20 03:52:05 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112619.M
Quant Title : GC Extractables
QLast Update : Wed Dec 18 14:58:42 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

