

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL121922\
 Data File : PL079773.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Dec 2022 16:02
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 19 20:35:04 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112322.M
 Quant Title : GC Extractables
 QLast Update : Thu Nov 24 00:09:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.321	2.561	109.1E6	132.9E6	48.295	48.815
28)	SA Decachlor...	8.768	7.629	81970863	99086845	42.207	42.178
Target Compounds							
2)	A alpha-BHC	3.782	3.048	169.3E6	212.0E6	49.236	49.161
3)	MA gamma-BHC...	4.115	3.369	158.8E6	190.7E6	49.006	48.975
4)	MA Heptachlor	4.685	3.695	158.4E6	190.2E6	48.080	49.444
5)	MB Aldrin	5.023	3.967	151.5E6	184.6E6	48.779	49.707
6)	B beta-BHC	4.331	3.670	66902004	72174503	47.382	47.634
7)	B delta-BHC	4.571	3.889	143.0E6	191.0E6	51.539	49.646
8)	B Heptachlo...	5.456	4.463	141.9E6	167.8E6	49.710	49.724
9)	A Endosulfan I	5.841	4.826	130.8E6	160.2E6	49.180	50.071
10)	B gamma-Chl...	5.711	4.711	141.5E6	170.1E6	49.465	50.249
11)	B alpha-Chl...	5.788	4.773	140.1E6	166.4E6	48.627	49.682
12)	B 4,4'-DDE	5.972	4.968	122.5E6	159.2E6	49.016	51.915
13)	MA Dieldrin	6.118	5.088	131.6E6	172.0E6	47.996	51.009
14)	MA Endrin	6.348	5.359	107.3E6	137.9E6	44.336	45.601
15)	B Endosulfa...	6.577	5.655	105.7E6	136.8E6	46.837	48.886
16)	A 4,4'-DDD	6.497	5.519	103.3E6	137.5E6	52.051	54.972
17)	MA 4,4'-DDT	6.805	5.765	95747080	121.1E6	42.210	46.339
18)	B Endrin al...	6.709	5.834	86301022	104.4E6	47.650	49.065
19)	B Endosulfa...	6.942	6.056	108.6E6	137.2E6	46.847	48.625
20)	A Methoxychlor	7.290	6.342	49243434	61472721	40.819	45.228
21)	B Endrin ke...	7.428	6.553	116.6E6	154.7E6	47.261	50.224
22)	Mirex	7.884	6.724	84479613	108.5E6	44.601	45.887

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL121922\
Data File : PL079773.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Dec 2022 16:02
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 19 20:35:04 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112322.M
Quant Title : GC Extractables
QLast Update : Thu Nov 24 00:09:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

