

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\
 Method File : LTX122022.M
 Title : GC Extractables
 Last Update : Wed Dec 21 03:42:34 2022
 Response Via : Initial Calibration

Calibration Files

500 =PL079806.D 1000 =PL079804.D 250 =PL079807.D
 750 =PL079805.D 100 =PL079808.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	2.120	2.168	1.996	2.066	1.917	2.054	E6	4.85
2)	Toxaphene-1	1.696	1.636	1.524	1.643	1.630	1.626	E4	3.84
3)	Toxaphene-2	6.178	6.237	6.294	6.108	5.916	6.146	E4	2.38
4)	Toxaphene-3	4.854	5.052	4.856	4.832	4.146	4.748	E4	7.33
5)	Toxaphene-4	3.454	3.541	3.307	3.329	3.052	3.337	E4	5.55
6)	Toxaphene-5	3.090	3.212	3.132	3.131	2.827	3.079	E4	4.79
7) SA	Decachlorobiphenyl	1.805	1.773	1.805	1.738	1.792	1.783	E6	1.57

Signal #2 Calibration Files

500 =PL079806.D 1000 =PL079804.D 250 =PL079807.D
 750 =PL079805.D 100 =PL079808.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	2.608	2.579	2.544	2.488	2.470	2.538	E6	2.31
2)	Toxaphene-1	1.684	1.557	1.574	1.561	1.597	1.595	E4	3.30
3)	Toxaphene-2	2.352	2.290	2.289	2.195	2.138	2.253	E4	3.78
4)	Toxaphene-3	6.941	6.969	6.529	6.645	5.711	6.559	E4	7.78
5)	Toxaphene-4	9.695	9.681	9.233	9.307	8.517	9.287	E4	5.16
6)	Toxaphene-5	6.082	5.953	5.728	5.810	4.964	5.707	E4	7.66
7) SA	Decachlorobiphenyl	2.265	2.205	2.270	2.158	2.326	2.245	E6	2.90

(#) = Out of Range