

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL122118\
 Data File : PL043349.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Dec 2018 12:48
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 12/26/2018 9:12:11 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 21 22:17:22 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL120618.M
 Quant Title : GC Extractables
 QLast Update : Fri Dec 07 00:20:08 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.465	4.100	318.0E6	169.3E6	51.854	57.109
28)	SA Decachlor...	8.204	9.234	283.4E6	129.1E6	43.401	49.489
Target Compounds							
2)	A alpha-BHC	3.905	4.496	493.7E6	261.3E6	52.220	57.764
3)	MA gamma-BHC...	4.188	4.785	446.2E6	238.8E6	51.680	56.947
4)	MA Heptachlor	4.483	5.300	437.2E6	220.6E6	49.270	53.842
5)	MB Aldrin	4.729	5.609	428.8E6	234.9E6	50.576	57.548
6)	B beta-BHC	4.442	4.954	182.7E6	97079300	49.543	57.344
7)	B delta-BHC	4.644	5.175	420.2E6	229.8E6	51.887	53.480
8)	B Heptachlo...	5.173	5.998	389.0E6	212.0E6	49.069	55.910
9)	A Endosulfan I	5.510	6.358	355.4E6	191.6E6	48.507	56.424
10)	B gamma-Chl...	5.399	6.233	395.3E6	209.1E6	49.029	56.933
11)	B alpha-Chl...	5.456	6.306	384.2E6	202.1E6	48.395	56.206
12)	B 4,4'-DDE	5.622	6.461	350.3E6	204.7E6	50.596	56.482
13)	MA Dieldrin	5.753	6.617	383.3E6	208.0E6	49.305	57.381
14)	MA Endrin	6.010	6.837	317.6E6	158.2E6	44.102	50.684
15)	B Endosulfa...	6.285	7.045	319.5E6	171.4E6	48.025	56.605
16)	A 4,4'-DDD	6.139	6.953	296.6E6	170.9E6	52.297	58.282
17)	MA 4,4'-DDT	6.377	7.254	263.7E6	145.1E6	44.321	49.390
18)	B Endrin al...	6.453	7.168	262.6E6	138.6E6	47.658	55.240m
19)	B Endosulfa...	6.667	7.393	320.6E6	162.2E6	47.857	53.946
20)	A Methoxychlor	6.923	7.710	142.7E6	67969903	42.514	45.063
21)	B Endrin ke...	7.155	7.868	364.3E6	167.0E6	50.378	54.642

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL122118\
Data File : PL043349.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Dec 2018 12:48
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
12/26/2018 9:12:11 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 21 22:17:22 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL120618.M
Quant Title : GC Extractables
QLast Update : Fri Dec 07 00:20:08 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

