

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL122118\
 Data File : PL043417.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Dec 2018 14:50
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 26 08:14:41 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL120618.M
 Quant Title : GC Extractables
 QLast Update : Fri Dec 07 00:20:08 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.465	4.102	346.6E6	170.9E6	56.507	57.655
28)	SA Decachlor...	8.202	9.235	303.9E6	140.3E6	46.545	53.800
Target Compounds							
2)	A alpha-BHC	3.904	4.497	536.1E6	252.4E6	56.712	55.800
3)	MA gamma-BHC...	4.187	4.786	486.2E6	233.1E6	56.304	55.575
4)	MA Heptachlor	4.483	5.301	480.5E6	221.4E6	54.158	54.054
5)	MB Aldrin	4.729	5.610	464.9E6	222.7E6	54.829	54.558
6)	B beta-BHC	4.442	4.955	199.0E6	100.8E6	53.956	59.556
7)	B delta-BHC	4.643	5.176	458.5E6	220.1E6	56.618	51.226
8)	B Heptachlor...	5.171	5.999	420.0E6	211.6E6	52.989	55.798
9)	A Endosulfan I	5.509	6.359	379.9E6	184.0E6	51.860	54.187
10)	B gamma-Chl...	5.397	6.233	426.6E6	203.3E6	52.910	55.352
11)	B alpha-Chl...	5.455	6.307	410.5E6	197.8E6	51.708	54.995
12)	B 4,4'-DDE	5.621	6.461	381.2E6	198.7E6	55.059	54.826
13)	MA Dieldrin	5.752	6.617	411.6E6	190.8E6	52.944	52.620
14)	MA Endrin	6.009	6.836	350.2E6	143.8E6	48.632	46.075
15)	B Endosulfa...	6.284	7.045	342.8E6	166.7E6	51.531	55.068
16)	A 4,4'-DDD	6.137	6.954	322.2E6	162.0E6	56.819	55.258
17)	MA 4,4'-DDT	6.375	7.254	286.7E6	140.8E6	48.189	47.947
18)	B Endrin al...	6.451	7.170	284.9E6	137.5E6	51.705	54.815
19)	B Endosulfa...	6.664	7.394	342.3E6	163.1E6	51.095	54.238
20)	A Methoxychlor	6.921	7.711	155.2E6	69256876	46.250	45.917
21)	B Endrin ke...	7.154	7.868	385.0E6	164.9E6	53.239	53.945

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL122118\
Data File : PL043417.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Dec 2018 14:50
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 26 08:14:41 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL120618.M
Quant Title : GC Extractables
QLast Update : Fri Dec 07 00:20:08 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

