

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL122119\
 Data File : PL055165.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Dec 2019 18:56
 Operator : SG\AJ
 Sample : K6315-03
 Misc : 608.3 METHOD
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 NCQ-632

Manual Integrations
 APPROVED

Ankita
 12/23/2019 12:08:05 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 21 05:33:04 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122119.M
 Quant Title : GC Extractables
 QLast Update : Sat Dec 21 05:21:06 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.532	4.049	372.6E6	52033891	35.426m	19.693 #
28) SA Decachlor...	8.309	9.144	172.8E6	40995324	18.628m	18.571
Target Compounds						
4) MA Heptachlor	4.569	5.233	67044360	43126142	4.106m	13.708 #
14) MA Endrin	6.109	6.757	68339087	6697775	5.377	3.121 #
17) MA 4,4'-DDT	6.472	7.180	28998969	2840661	2.881m	1.447 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL122119\
Data File : PL055165.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Dec 2019 18:56
Operator : SG\AJ
Sample : K6315-03
Misc : 608.3 METHOD
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
NCQ-632

Manual Integrations
APPROVED

Ankita
12/23/2019 12:08:05 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 21 05:33:04 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122119.M
Quant Title : GC Extractables
QLast Update : Sat Dec 21 05:21:06 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

