

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL122122\
 Data File : PL079840.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Dec 2022 16:08
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 21 21:45:20 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122022.M
 Quant Title : GC Extractables
 QLast Update : Wed Dec 21 10:00:39 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.317	2.562	112.0E6	134.0E6	50.983	50.481
28)	SA Decachlor...	8.762	7.627	92226649	108.1E6	49.725	45.032
Target Compounds							
2)	A alpha-BHC	3.777	3.049	172.7E6	214.7E6	51.218	49.680
3)	MA gamma-BHC...	4.110	3.370	161.5E6	193.5E6	50.559	49.403
4)	MA Heptachlor	4.680	3.696	163.1E6	191.4E6	49.998	47.150
5)	MB Aldrin	5.017	3.967	149.5E6	180.4E6	49.562	47.065
6)	B beta-BHC	4.326	3.670	69123309	72101006	50.458	47.544
7)	B delta-BHC	4.566	3.890	142.8E6	191.9E6	53.185	48.216
8)	B Heptachlo...	5.451	4.464	137.5E6	164.0E6	51.937	46.994
9)	A Endosulfan I	5.835	4.825	128.1E6	154.6E6	49.328	46.141
10)	B gamma-Chl...	5.706	4.711	138.0E6	164.1E6	49.700	46.968
11)	B alpha-Chl...	5.783	4.773	136.7E6	160.5E6	49.136	46.738
12)	B 4,4'-DDE	5.967	4.967	120.2E6	152.1E6	49.841	47.298
13)	MA Dieldrin	6.113	5.087	130.7E6	166.1E6	49.523	48.006
14)	MA Endrin	6.343	5.358	113.7E6	140.3E6	48.618	49.363
15)	B Endosulfa...	6.573	5.655	108.7E6	132.3E6	44.596	44.907
16)	A 4,4'-DDD	6.491	5.518	99494472	126.4E6	50.730	49.616
17)	MA 4,4'-DDT	6.800	5.765	105.7E6	125.0E6	47.235	46.146
18)	B Endrin al...	6.705	5.833	87764981	103.1E6	49.332	45.949
19)	B Endosulfa...	6.937	6.056	111.6E6	134.4E6	48.634	46.297
20)	A Methoxychlor	7.285	6.342	55415628	65815361	47.013	46.412
21)	B Endrin ke...	7.424	6.552	119.2E6	148.1E6	48.835	46.379
22)	Mirex	7.880	6.723	90463454	108.6E6	47.206	44.451

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL122122\
Data File : PL079840.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Dec 2022 16:08
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 21 21:45:20 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122022.M
Quant Title : GC Extractables
QLast Update : Wed Dec 21 10:00:39 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

