

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL122220\
 Data File : PL063741.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Dec 2020 09:33
 Operator : AR\AJ
 Sample : PSTDICC100
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC100

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 22 10:43:11 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122220.M
 Quant Title : GC Extractables
 QLast Update : Tue Dec 22 10:40:33 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.440	4.113	300.4E6	254.5E6	98.197	97.697
28)	SA Decachlor...	8.183	9.274	248.6E6	180.0E6	92.030	92.059
Target Compounds							
2)	A alpha-BHC	3.875	4.515	516.4E6	386.8E6	101.204	102.495
3)	MA gamma-BHC...	4.156	4.806	478.8E6	357.3E6	101.087	100.978
4)	MA Heptachlor	4.457	5.319	426.5E6	317.4E6	99.657	98.958
5)	MB Aldrin	4.703	5.629	412.5E6	294.6E6	100.436	100.935
6)	B beta-BHC	4.405	4.981	184.1E6	152.0E6	96.326	93.197
7)	B delta-BHC	4.608	5.203	465.0E6	338.6E6	99.444	99.366
8)	B Heptachlo...	5.144	6.023	377.1E6	275.2E6	98.788	98.098
9)	A Endosulfan I	5.481	6.385	346.2E6	250.0E6	98.363	98.204
10)	B gamma-Chl...	5.371	6.258	383.3E6	276.5E6	99.229	98.829
11)	B alpha-Chl...	5.429	6.331	375.8E6	274.0E6	98.797	98.183
12)	B 4,4'-DDE	5.596	6.487	356.3E6	254.4E6	101.191	101.016
13)	MA Dieldrin	5.724	6.646	372.0E6	264.3E6	100.222	99.953
14)	MA Endrin	5.978	6.868	299.7E6	199.4E6	101.749	101.814
15)	B Endosulfa...	6.251	7.081	322.4E6	228.8E6	97.399	97.127
16)	A 4,4'-DDD	6.110	6.985	297.6E6	213.2E6	98.963	98.832
17)	MA 4,4'-DDT	6.350	7.285	281.5E6	200.4E6	102.838	101.558
18)	B Endrin al...	6.419	7.206	266.4E6	193.4E6	96.033	95.476
19)	B Endosulfa...	6.633	7.430	308.7E6	215.1E6	96.606	95.563
20)	A Methoxychlor	6.900	7.744	129.8E6	102.6E6	97.895	96.326
21)	B Endrin ke...	7.120	7.910	343.8E6	240.2E6	96.576	95.815
22)	Mirex	7.308	8.362	236.0E6	165.8E6	93.954	92.612

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL122220\
Data File : PL063741.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Dec 2020 09:33
Operator : AR\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 22 10:43:11 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122220.M
Quant Title : GC Extractables
QLast Update : Tue Dec 22 10:40:33 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

